

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022323.D
 Acq On : 02 Jun 2021 04:45
 Operator : JC/MD
 Sample : M2572-11 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20210528

Quant Time: Jun 02 05:35:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	63064	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	133994	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120112	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44905	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54256	55.77	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.54%
35) Dibromofluoromethane	5.391	113	40469	47.63	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.26%
50) Toluene-d8	8.653	98	168394	51.45	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.90%
62) 4-Bromofluorobenzene	11.085	95	57778	47.76	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.52%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	537	0.70	ug/l	95
4) Vinyl Chloride	1.374	62	4295	4.91	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.337	101	435	0.58	ug/l	# 74
12) 1,1-Dichloroethene	2.319	96	550	0.76	ug/l	93
21) trans-1,2-Dichloroethene	3.105	96	554	0.70	ug/l	# 62
27) cis-1,2-Dichloroethene	4.501	96	64744	67.72	ug/l	91
44) Trichloroethene	7.135	130	115988	122.43	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022323.D
Acq On : 02 Jun 2021 04:45
Operator : JC/MD
Sample : M2572-11 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20210528

Quant Time: Jun 02 05:35:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

