

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060122\  
 Data File : VX029108.D  
 Acq On : 01 Jun 2022 11:10  
 Operator : JC/MD  
 Sample : VX0601WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_X  
**ClientSampleId :**  
 VX0601WBL01

Quant Time: Jun 02 04:42:34 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X053122W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 01 04:45:42 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.556	168	328702	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	569151	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	542321	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	254677	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	189281	51.066	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.140%
35) Dibromofluoromethane	5.391	113	187952	49.717	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.440%
50) Toluene-d8	8.653	98	660287	47.755	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.500%
62) 4-Bromofluorobenzene	11.079	95	247427	46.294	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.580%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060122\  
 Data File : VX029108.D  
 Acq On : 01 Jun 2022 11:10  
 Operator : JC/MD  
 Sample : VX0601WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0601WBL01

Quant Time: Jun 02 04:42:34 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X053122W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 01 04:45:42 2022  
 Response via : Initial Calibration

