

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
 Data File : VX029112.D
 Acq On : 01 Jun 2022 12:50
 Operator : JC/MD
 Sample : N3022-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22BR-20220525

Quant Time: Jun 02 04:44:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	344383	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	591919	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	575227	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	279010	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	194089	49.979	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.960%
35) Dibromofluoromethane	5.385	113	197377	50.202	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.400%
50) Toluene-d8	8.653	98	692867	48.184	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.360%
62) 4-Bromofluorobenzene	11.079	95	271401	48.826	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.660%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3238	2.117	ug/l #	82
27) cis-1,2-Dichloroethene	4.489	96	18453	4.419	ug/l	86
30) Chloroform	5.093	83	2423	0.379	ug/l	90
44) Trichloroethene	7.123	130	13305	2.326	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
Data File : VX029112.D
Acq On : 01 Jun 2022 12:50
Operator : JC/MD
Sample : N3022-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22BR-20220525

Quant Time: Jun 02 04:44:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

