

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
 Data File : VX029114.D
 Acq On : 01 Jun 2022 13:37
 Operator : JC/MD
 Sample : N3022-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB06-20220525

Quant Time: Jun 02 04:45:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	327989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	570861	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	561788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	267466	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	189450	51.223	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.440%	
35) Dibromofluoromethane	5.391	113	190166	50.152	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.300%	
50) Toluene-d8	8.653	98	671669	48.433	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.860%	
62) 4-Bromofluorobenzene	11.079	95	259185	48.348	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.700%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	19966	13.709	ug/l	100
20) Methylene Chloride	2.788	84	2661	0.784	ug/l	96
25) 2-Butanone	4.574	43	3672	1.563	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
Data File : VX029114.D
Acq On : 01 Jun 2022 13:37
Operator : JC/MD
Sample : N3022-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB06-20220525

Quant Time: Jun 02 04:45:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

