

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
 Data File : VX029116.D
 Acq On : 01 Jun 2022 14:24
 Operator : JC/MD
 Sample : N3022-09DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 91BR-20220525DL

Quant Time: Jun 02 04:46:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	328888	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	564222	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	552279	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	269065	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	187136	50.459	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.920%	
35) Dibromofluoromethane	5.391	113	188108	50.193	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.380%	
50) Toluene-d8	8.653	98	657412	47.962	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.920%	
62) 4-Bromofluorobenzene	11.085	95	257241	48.550	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.100%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1058	0.370	ug/l	82
4) Vinyl Chloride	1.373	62	55113	16.749	ug/l	97
21) trans-1,2-Dichloroethene	3.099	96	2582	0.772	ug/l #	77
27) cis-1,2-Dichloroethene	4.489	96	177590	44.529	ug/l	83
44) Trichloroethene	7.129	130	9809	1.637	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
Data File : VX029116.D
Acq On : 01 Jun 2022 14:24
Operator : JC/MD
Sample : N3022-09DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
91BR-20220525DL

Quant Time: Jun 02 04:46:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

