

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
 Data File : VX029123.D
 Acq On : 01 Jun 2022 17:07
 Operator : JC/MD
 Sample : N3022-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 25BR-20220525

Quant Time: Jun 02 04:49:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	324589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	558963	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	550469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	262406	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	183652	50.175	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.360%	
35) Dibromofluoromethane	5.391	113	185451	49.950	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.900%	
50) Toluene-d8	8.653	98	657421	48.414	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.820%	
62) 4-Bromofluorobenzene	11.079	95	250660	47.753	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.500%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	823	0.292	ug/l	91
4) Vinyl Chloride	1.374	62	10661	3.283	ug/l	100
21) trans-1,2-Dichloroethene	3.099	96	1454	0.441	ug/l #	71
27) cis-1,2-Dichloroethene	4.489	96	234576	59.597	ug/l	82
44) Trichloroethene	7.129	130	428651	102.935	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
Data File : VX029123.D
Acq On : 01 Jun 2022 17:07
Operator : JC/MD
Sample : N3022-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
25BR-20220525

Quant Time: Jun 02 04:49:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

