

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
 Data File : VX029124.D
 Acq On : 01 Jun 2022 17:31
 Operator : JC/MD
 Sample : N3022-05 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 56BR-20220525

Quant Time: Jun 02 09:11:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	336901	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	582362	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	571200	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	280857	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	189158	49.791	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.580%
35) Dibromofluoromethane	5.391	113	192239	49.698	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.400%
50) Toluene-d8	8.653	98	679839	48.054	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.100%
62) 4-Bromofluorobenzene	11.085	95	266489	48.729	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.460%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	3075	0.912	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	62218	15.230	ug/l	82
44) Trichloroethene	7.129	130	334109	76.829	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
Data File : VX029124.D
Acq On : 01 Jun 2022 17:31
Operator : JC/MD
Sample : N3022-05 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
56BR-20220525

Quant Time: Jun 02 09:11:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

