

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
 Data File : VX029130.D
 Acq On : 01 Jun 2022 19:51
 Operator : JC/MD
 Sample : N3087-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Jun 02 09:14:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	366657	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	621765	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	605439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	302974	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	187739	45.407	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.820%
35) Dibromofluoromethane	5.391	113	199304	48.259	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.520%
50) Toluene-d8	8.653	98	721569	47.771	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.540%
62) 4-Bromofluorobenzene	11.079	95	285411	48.882	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.760%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	4127	2.535	ug/l	88
20) Methylene Chloride	2.788	84	1229	0.324	ug/l #	81
84) 1,2,4-Trimethylbenzene	11.750	105	4796	0.276	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
Data File : VX029130.D
Acq On : 01 Jun 2022 19:51
Operator : JC/MD
Sample : N3087-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Jun 02 09:14:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

