

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\X060122\
 Data File : VX029106.D
 Acq On : 01 Jun 2022 10:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 02 04:41:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	50.000	56.584	-13.2	100	0.00
3 P	Chloromethane	50.000	53.027	-6.1	102	0.00
4 C	Vinyl Chloride	50.000	54.552	-9.1#	99	0.00
5 T	Bromomethane	50.000	55.092	-10.2	97	0.00
6 T	Chloroethane	50.000	48.212	3.6	96	0.00
7 T	Trichlorofluoromethane	50.000	54.754	-9.5	101	0.00
8 T	Diethyl Ether	50.000	54.075	-8.2	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	55.509	-11.0	105	0.00
10 T	Methyl Iodide	50.000	55.353	-10.7	102	0.00
11 T	Tert butyl alcohol	250.000	223.507	10.6	86	0.00
12 CM	1,1-Dichloroethene	50.000	53.070	-6.1#	100	0.00
13 T	Acrolein	250.000	213.115	14.8	79	0.00
14 T	Allyl chloride	50.000	52.029	-4.1	96	0.00
15 T	Acrylonitrile	250.000	263.553	-5.4	97	0.00
16 T	Acetone	250.000	291.483	-16.6	110	0.00
17 T	Carbon Disulfide	50.000	48.492	3.0	93	0.00
18 T	Methyl Acetate	50.000	51.995	-4.0	99	0.00
19 T	Methyl tert-butyl Ether	50.000	52.510	-5.0	96	0.00
20 T	Methylene Chloride	50.000	52.941	-5.9	100	0.00
21 T	trans-1,2-Dichloroethene	50.000	52.995	-6.0	99	0.00
22 T	Diisopropyl ether	50.000	54.453	-8.9	99	0.00
23 T	Vinyl Acetate	250.000	274.025	-9.6	97	0.00
24 P	1,1-Dichloroethane	50.000	53.886	-7.8	100	0.00
25 T	2-Butanone	250.000	269.516	-7.8	100	0.00
26 T	2,2-Dichloropropane	50.000	51.887	-3.8	96	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.971	-3.9	99	0.00
28 T	Bromochloromethane	50.000	51.169	-2.3	101	0.00
29 T	Tetrahydrofuran	250.000	262.769	-5.1	95	0.00
30 C	Chloroform	50.000	53.857	-7.7#	99	0.00
31 T	Cyclohexane	50.000	55.275	-10.5	101	0.00
32 T	1,1,1-Trichloroethane	50.000	52.799	-5.6	96	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.093	-0.2	94	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	92	0.00
35 S	Dibromofluoromethane	50.000	52.046	-4.1	94	0.00
36 T	1,1-Dichloropropene	50.000	56.246	-12.5	101	0.00
37 T	Ethyl Acetate	50.000	54.063	-8.1	93	0.00
38 T	Carbon Tetrachloride	50.000	54.540	-9.1	96	0.00
39 T	Methylcyclohexane	50.000	58.566	-17.1	103	0.00
40 TM	Benzene	50.000	56.458	-12.9	100	0.00
41 T	Methacrylonitrile	50.000	56.531	-13.1	95	0.00
42 TM	1,2-Dichloroethane	50.000	55.420	-10.8	98	0.00
43 T	Isopropyl Acetate	50.000	53.397	-6.8	93	0.00
44 TM	Trichloroethene	50.000	56.931	-13.9	100	0.00
45 C	1,2-Dichloropropane	50.000	57.097	-14.2#	101	0.00
46 T	Dibromomethane	50.000	55.465	-10.9	99	0.00
47 T	Bromodichloromethane	50.000	54.030	-8.1	94	0.00
48 T	Methyl methacrylate	50.000	54.913	-9.8	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\X060122\
 Data File : VX029106.D
 Acq On : 01 Jun 2022 10:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 02 04:41:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1081.668	-8.2	96	0.00
50 S	Toluene-d8	50.000	52.927	-5.9	95	0.00
51 T	4-Methyl-2-Pentanone	250.000	274.643	-9.9	96	0.00
52 CM	Toluene	50.000	56.961	-13.9#	100	0.00
53 T	t-1,3-Dichloropropene	50.000	53.443	-6.9	91	0.00
54 T	cis-1,3-Dichloropropene	50.000	55.558	-11.1	96	0.00
55 T	1,1,2-Trichloroethane	50.000	55.656	-11.3	99	0.00
56 T	Ethyl methacrylate	50.000	56.103	-12.2	96	0.00
57 T	1,3-Dichloropropane	50.000	56.203	-12.4	98	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	254.234	-1.7	85	0.00
59 T	2-Hexanone	250.000	280.616	-12.2	97	0.00
60 T	Dibromochloromethane	50.000	53.503	-7.0	93	0.00
61 T	1,2-Dibromoethane	50.000	55.350	-10.7	99	0.00
62 S	4-Bromofluorobenzene	50.000	52.735	-5.5	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	93	0.00
64 T	Tetrachloroethene	50.000	57.105	-14.2	104	0.00
65 PM	Chlorobenzene	50.000	55.202	-10.4	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.562	-7.1	97	0.00
67 C	Ethyl Benzene	50.000	56.741	-13.5#	99	0.00
68 T	m/p-Xylenes	100.000	115.145	-15.1	100	0.00
69 T	o-Xylene	50.000	56.909	-13.8	99	0.00
70 T	Styrene	50.000	58.114	-16.2	100	0.00
71 P	Bromoform	50.000	49.447	1.1	88	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	94	0.00
73 T	Isopropylbenzene	50.000	54.893	-9.8	100	0.00
74 T	N-ethyl acetate	50.000	54.048	-8.1	95	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	53.076	-6.2	98	0.00
76 T	1,2,3-Trichloropropane	50.000	53.855	-7.7	100	0.00
77 T	Bromobenzene	50.000	52.387	-4.8	99	0.00
78 T	n-propylbenzene	50.000	56.838	-13.7	102	0.00
79 T	2-Chlorotoluene	50.000	53.840	-7.7	100	0.00
80 T	1,3,5-Trimethylbenzene	50.000	55.865	-11.7	101	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.556	0.9	89	0.00
82 T	4-Chlorotoluene	50.000	55.054	-10.1	101	0.00
83 T	tert-Butylbenzene	50.000	54.280	-8.6	99	0.00
84 T	1,2,4-Trimethylbenzene	50.000	55.848	-11.7	101	0.00
85 T	sec-Butylbenzene	50.000	56.921	-13.8	102	0.00
86 T	p-Isopropyltoluene	50.000	57.258	-14.5	103	0.00
87 T	1,3-Dichlorobenzene	50.000	53.963	-7.9	100	0.00
88 T	1,4-Dichlorobenzene	50.000	53.915	-7.8	102	0.00
89 T	n-Butylbenzene	50.000	58.406	-16.8	104	0.00
90 T	Hexachloroethane	50.000	50.377	-0.8	91	0.00
91 T	1,2-Dichlorobenzene	50.000	55.966	-11.9	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.358	-2.7	92	0.00
93 T	1,2,4-Trichlorobenzene	50.000	57.267	-14.5	107	0.00
94 T	Hexachlorobutadiene	50.000	58.946	-17.9	105	0.00
95 T	Naphthalene	50.000	57.653	-15.3	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060122\
Data File : VX029106.D
Acq On : 01 Jun 2022 10:14
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 02 04:41:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	57.302	-14.6	105	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6