

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060123\
 Data File : VX035957.D
 Acq On : 01 Jun 2023 14:14
 Operator : JC/MD
 Sample : 03002-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-48

Quant Time: Jun 02 01:41:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	218411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	372246	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	326198	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	151964	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	163549	46.113	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.220%		
35) Dibromofluoromethane	5.379	113	113118	45.180	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.360%		
50) Toluene-d8	8.647	98	431248	47.993	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	95.980%		
62) 4-Bromofluorobenzene	11.079	95	164372	46.339	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.680%		
Target Compounds						
16) Acetone	2.392	43	3816	2.608	ug/l	99
17) Carbon Disulfide	2.514	76	3070	0.559	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060123\
Data File : VX035957.D
Acq On : 01 Jun 2023 14:14
Operator : JC/MD
Sample : 03002-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-48

Quant Time: Jun 02 01:41:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

