

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060123\
 Data File : VX035962.D
 Acq On : 01 Jun 2023 16:11
 Operator : JC/MD
 Sample : 03005-02 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-12

Quant Time: Jun 02 01:42:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	203214	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	328910	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	287971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135612	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	141220	42.795	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.600%
35) Dibromofluoromethane	5.385	113	100348	45.361	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.720%
50) Toluene-d8	8.647	98	385880	48.602	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.200%
62) 4-Bromofluorobenzene	11.079	95	146631	46.784	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060123\
Data File : VX035962.D
Acq On : 01 Jun 2023 16:11
Operator : JC/MD
Sample : 03005-02 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-12

Quant Time: Jun 02 01:42:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

