

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060123\
 Data File : VX035965.D
 Acq On : 01 Jun 2023 17:22
 Operator : JC/MD
 Sample : 03005-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-11

Quant Time: Jun 02 01:42:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	198344	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	329669	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	292280	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134625	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	139911	43.440	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.880%
35) Dibromofluoromethane	5.385	113	99512	44.879	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.760%
50) Toluene-d8	8.647	98	386934	48.623	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.240%
62) 4-Bromofluorobenzene	11.079	95	147204	46.859	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.720%
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.520	76	1167	0.234	ug/l #	74
19) Methyl tert-butyl Ether	3.117	73	6519	0.836	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060123\
Data File : VX035965.D
Acq On : 01 Jun 2023 17:22
Operator : JC/MD
Sample : 03005-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-11

Quant Time: Jun 02 01:42:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

