

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060123\
 Data File : VX035966.D
 Acq On : 01 Jun 2023 17:45
 Operator : JC/MD
 Sample : 03005-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-12

Quant Time: Jun 02 01:42:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	187218	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	308601	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	294106	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135296	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	145828	47.967	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.940%
35) Dibromofluoromethane	5.391	113	95829	46.169	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.340%
50) Toluene-d8	8.647	98	381379	51.197	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.400%
62) 4-Bromofluorobenzene	11.079	95	148149	50.379	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.760%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2916	2.325	ug/l #	87
19) Methyl tert-butyl Ether	3.111	73	2447	0.333	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060123\
Data File : VX035966.D
Acq On : 01 Jun 2023 17:45
Operator : JC/MD
Sample : 03005-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-12

Quant Time: Jun 02 01:42:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

