



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060123\
Data File : VX035945.D
Acq On : 01 Jun 2023 09:00
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Tue May 23 11:36:37 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	13939	PASS
75	95	30	60	50.0	37184	PASS
95	95	100	100	100.0	74307	PASS
96	95	5	9	6.4	4759	PASS
173	174	0.00	2	0.9	420	PASS
174	95	50	100	64.5	47904	PASS
175	174	5	9	7.4	3537	PASS
176	174	95	101	96.9	46397	PASS
177	176	5	9	6.8	3158	PASS