

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022328.D
 Acq On : 02 Jun 2021 10:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 03 02:17:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	90169	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	175096	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	161547	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73514	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	74299	53.42	ug/l	0.00
Spiked Amount 50.000	Range 78	- 117	Recovery	=	106.84%	
35) Dibromofluoromethane	5.391	113	55262	49.77	ug/l	0.00
Spiked Amount 50.000	Range 75	- 124	Recovery	=	99.54%	
50) Toluene-d8	8.653	98	218914	51.18	ug/l	0.00
Spiked Amount 50.000	Range 92	- 112	Recovery	=	102.36%	
62) 4-Bromofluorobenzene	11.085	95	84315	53.33	ug/l	0.00
Spiked Amount 50.000	Range 83	- 123	Recovery	=	106.66%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.172	85	57009	52.01	ug/l	99
3) Chloromethane	1.294	50	62867	53.17	ug/l	99
4) Vinyl Chloride	1.373	62	64313	51.40	ug/l	99
5) Bromomethane	1.599	94	36447	49.96	ug/l	98
6) Chloroethane	1.678	64	38817	52.10	ug/l	99
7) Trichlorofluoromethane	1.886	101	93495	51.97	ug/l	94
8) Diethyl Ether	2.136	74	37883	48.96	ug/l	88
9) 1,1,2-Trichlorotrifluo...	2.331	101	56575	52.96	ug/l	91
10) Methyl Iodide	2.453	142	63340	55.17	ug/l	95
11) Tert butyl alcohol	2.977	59	83559	262.66	ug/l	99
12) 1,1-Dichloroethene	2.318	96	55246	53.46	ug/l	98
13) Acrolein	2.239	56	55945	265.15	ug/l	98
14) Allyl chloride	2.666	41	105295	55.47	ug/l	96
15) Acrylonitrile	3.068	53	177355	283.17	ug/l	99
16) Acetone	2.385	43	180744	298.43	ug/l	96
17) Carbon Disulfide	2.514	76	144751	53.48	ug/l	100
18) Methyl Acetate	2.709	43	88372	54.87	ug/l	93
19) Methyl tert-butyl Ether	3.123	73	191180	54.00	ug/l	99
20) Methylene Chloride	2.794	84	66963	49.80	ug/l	93
21) trans-1,2-Dichloroethene	3.093	96	62466	55.57	ug/l	95
22) Diisopropyl ether	3.769	45	204592	55.96	ug/l #	95
23) Vinyl Acetate	3.727	43	1015989	288.64	ug/l	95
24) 1,1-Dichloroethane	3.617	63	117713	54.87	ug/l	98
25) 2-Butanone	4.568	43	290072	296.05	ug/l	94
26) 2,2-Dichloropropane	4.483	77	98697	51.82	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	73533	53.79	ug/l	98
28) Bromochloromethane	4.903	49	56531	54.32	ug/l	83
29) Tetrahydrofuran	5.013	42	163722	280.74	ug/l	90
30) Chloroform	5.098	83	116309	52.47	ug/l	99
31) Cyclohexane	5.470	56	99132	56.72	ug/l	95
32) 1,1,1-Trichloroethane	5.391	97	99735	52.65	ug/l	97
36) 1,1-Dichloropropene	5.696	75	89440	55.24	ug/l	97
37) Ethyl Acetate	4.720	43	111101	55.98	ug/l	94
38) Carbon Tetrachloride	5.684	117	82145	51.86	ug/l	95
39) Methylcyclohexane	7.385	83	102897	56.03	ug/l	96
40) Benzene	6.043	78	272142	54.63	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022328.D
 Acq On : 02 Jun 2021 10:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 03 02:17:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	59288	55.84	ug/l	95
42) 1,2-Dichloroethane	6.092	62	95250	54.58	ug/l	98
43) Isopropyl Acetate	6.348	43	184354	56.79	ug/l	93
44) Trichloroethene	7.129	130	65064	52.56	ug/l	92
45) 1,2-Dichloropropane	7.433	63	71864	56.55	ug/l	98
46) Dibromomethane	7.586	93	47499	55.09	ug/l #	85
47) Bromodichloromethane	7.824	83	89947	52.14	ug/l	97
48) Methyl methacrylate	7.696	41	85797	57.35	ug/l	90
49) 1,4-Dioxane	7.659	88	32575	1118.28	ug/l	95
51) 4-Methyl-2-Pentanone	8.579	43	567220	291.68	ug/l	91
52) Toluene	8.720	92	169623	53.77	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	107578	54.02	ug/l	97
54) cis-1,3-Dichloropropene	8.372	75	118316	55.26	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	68744	53.68	ug/l	97
56) Ethyl methacrylate	9.122	69	115668	55.88	ug/l #	87
57) 1,3-Dichloropropane	9.311	76	121127	55.39	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	279020	249.24	ug/l	96
59) 2-Hexanone	9.433	43	436406	298.38	ug/l	90
60) Dibromochloromethane	9.524	129	64579	51.83	ug/l	99
61) 1,2-Dibromoethane	9.616	107	71000	55.68	ug/l	100
64) Tetrachloroethene	9.274	164	56254	53.04	ug/l #	88
65) Chlorobenzene	10.079	112	174575	51.91	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	60794	52.03	ug/l	97
67) Ethyl Benzene	10.195	91	328827	53.15	ug/l	100
68) m/p-Xylenes	10.305	106	246759	105.54	ug/l	99
69) o-Xylene	10.646	106	118662	52.18	ug/l	97
70) Styrene	10.658	104	203567	54.65	ug/l	98
71) Bromoform	10.805	173	40182	45.02	ug/l #	98
73) Isopropylbenzene	10.963	105	316455	50.86	ug/l	99
74) N-amyl acetate	10.847	43	168483	57.05	ug/l	90
75) 1,1,2,2-Tetrachloroethane	11.213	83	110898	53.77	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	128889	71.07	ug/l	90
77) Bromobenzene	11.201	156	66704	50.56	ug/l	85
78) n-propylbenzene	11.305	91	384949	52.82	ug/l	97
79) 2-Chlorotoluene	11.366	91	223381	50.54	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	266082	51.60	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.024	75	37828	49.42	ug/l	94
82) 4-Chlorotoluene	11.457	91	260220	51.10	ug/l	97
83) tert-Butylbenzene	11.719	119	247727	50.09	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	271433	52.29	ug/l	98
85) sec-Butylbenzene	11.896	105	329084	52.36	ug/l	97
86) p-Isopropyltoluene	12.012	119	271496	51.15	ug/l	96
87) 1,3-Dichlorobenzene	11.969	146	131513	51.70	ug/l	96
88) 1,4-Dichlorobenzene	12.042	146	130611	49.68	ug/l	97
89) n-Butylbenzene	12.335	91	265328	54.04	ug/l	98
90) Hexachloroethane	12.542	117	42290	48.21	ug/l	71
91) 1,2-Dichlorobenzene	12.341	146	123355	50.49	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.945	75	23326	52.86	ug/l	81
93) 1,2,4-Trichlorobenzene	13.591	180	76455	51.86	ug/l	96
94) Hexachlorobutadiene	13.725	225	26245	45.61	ug/l	97
95) Naphthalene	13.780	128	311454	54.17	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	75498	50.80	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
Data File : VX022328.D
Acq On : 02 Jun 2021 10:35
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jun 03 02:17:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022328.D
 Acq On : 02 Jun 2021 10:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 03 02:17:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.556 min Scan# 733
Delta R.T. -0.006 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:168 Resp: 90169
Ion Ratio Lower Upper
168 100
99 68.1 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 52.01 ug/l
RT: 1.172 min Scan# 14
Delta R.T. 0.005 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 85 Resp: 57009
Ion Ratio Lower Upper
85 100
87 32.4 16.4 49.2









#3
Chloromethane
Concen: 53.17 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 50 Resp: 62867
Ion Ratio Lower Upper
50 100
52 32.4 26.4 39.6



#4
Vinyl Chloride
Concen: 51.40 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 62 Resp: 64313
Ion Ratio Lower Upper
62 100
64 32.8 25.7 38.5









#5
Bromomethane
Concen: 49.96 ug/l
RT: 1.599 min Scan# 84
Delta R.T. -0.013 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 94 Resp: 36447
Ion Ratio Lower Upper
94 100
96 93.7 76.4 114.6



#6
Chloroethane
Concen: 52.10 ug/l
RT: 1.678 min Scan# 97
Delta R.T. -0.007 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 64 Resp: 38817
Ion Ratio Lower Upper
64 100
66 31.1 25.4 38.0









#7
Trichlorofluoromethane
Concen: 51.97 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 93495
Ion Ratio Lower Upper
101 100
103 61.7 53.2 79.8



#8
Diethyl Ether
Concen: 48.96 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 74 Resp: 37883
Ion Ratio Lower Upper
74 100
45 100.6 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.96 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 56575
Ion Ratio Lower Upper
101 100
85 45.8 34.9 52.3
151 66.8 61.4 92.2



#10
Methyl Iodide
Concen: 55.17 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion:142 Resp: 63340
Ion Ratio Lower Upper
142 100
127 45.4 33.3 49.9
141 14.1 11.4 17.0









#11
Tert butyl alcohol
Concen: 262.66 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.012 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 59 Resp: 83559
Ion Ratio Lower Upper
59 100
57 10.1 7.8 11.8



#12
1,1-Dichloroethene
Concen: 53.46 ug/l
RT: 2.318 min Scan# 202
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 96 Resp: 55246
Ion Ratio Lower Upper
96 100
61 159.8 124.6 187.0
98 63.7 50.6 76.0









#13
Acrolein
Concen: 265.15 ug/l
RT: 2.239 min Scan# 189
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 56 Resp: 55945
Ion Ratio Lower Upper
56 100
55 71.9 56.4 84.6



#14
Allyl chloride
Concen: 55.47 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 41 Resp: 105295
Ion Ratio Lower Upper
41 100
39 66.9 51.6 77.4
76 34.2 30.5 45.7









#15
Acrylonitrile
Concen: 283.17 ug/l
RT: 3.068 min Scan# 325
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 53 Resp: 177355
Ion Ratio Lower Upper
53 100
52 82.9 65.3 97.9
51 35.2 28.2 42.2



#16
Acetone
Concen: 298.43 ug/l
RT: 2.385 min Scan# 213
Delta R.T. -0.007 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 43 Resp: 180744
Ion Ratio Lower Upper
43 100
58 32.3 27.8 41.8









#17
Carbon Disulfide
Concen: 53.48 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 144751
Ion Ratio Lower Upper
76 100
78 9.1 7.2 10.8



#18
Methyl Acetate
Concen: 54.87 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 43 Resp: 88372
Ion Ratio Lower Upper
43 100
74 24.0 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 54.00 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.006 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 73 Resp: 191180
Ion Ratio Lower Upper
73 100
57 21.9 17.1 25.7



#20
Methylene Chloride
Concen: 49.80 ug/l
RT: 2.794 min Scan# 280
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 84 Resp: 66963
Ion Ratio Lower Upper
84 100
49 123.0 88.6 132.8
51 37.8 27.0 40.6
86 62.9 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 55.57 ug/l
RT: 3.093 min Scan# 329
Delta R.T. -0.006 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 62466
Ion Ratio Lower Upper
96 100
61 143.2 107.7 161.5
98 63.3 50.4 75.6



#22
Diisopropyl ether
Concen: 55.96 ug/l
RT: 3.769 min Scan# 440
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 45 Resp: 204592
Ion Ratio Lower Upper
45 100
43 70.8 58.0 87.0
87 25.4 25.4 38.0#
59 10.9 10.2 15.2









#23
 Vinyl Acetate
 Concen: 288.64 ug/l
 RT: 3.727 min Scan# 433
 Delta R.T. -0.000 min
 Lab File: VX022328.D
 Acq: 02 Jun 2021 10:35




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 43 Resp: 1015989
 Ion Ratio Lower Upper
 43 100
 86 10.4 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 54.87 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. -0.000 min
 Lab File: VX022328.D
 Acq: 02 Jun 2021 10:35

Tgt Ion: 63 Resp: 117713
 Ion Ratio Lower Upper
 63 100
 98 7.4 3.8 11.2
 100 3.8 2.5 7.5









#25
2-Butanone
Concen: 296.05 ug/l
RT: 4.568 min Scan# 571
Delta R.T. 0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 290072
Ion Ratio Lower Upper
43 100
72 26.4 23.7 35.5



#26
2,2-Dichloropropane
Concen: 51.82 ug/l
RT: 4.483 min Scan# 557
Delta R.T. 0.006 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 77 Resp: 98697
Ion Ratio Lower Upper
77 100
97 22.3 11.7 35.1









#27
cis-1,2-Dichloroethene
Concen: 53.79 ug/l
RT: 4.489 min Scan# 558
Delta R.T. -0.006 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 96 Resp: 73533
Ion Ratio Lower Upper
96 100
61 146.0 0.0 285.6
98 64.3 0.0 129.6



#28
Bromochloromethane
Concen: 54.32 ug/l
RT: 4.903 min Scan# 626
Delta R.T. -0.007 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 49 Resp: 56531
Ion Ratio Lower Upper
49 100
129 1.7 0.0 4.0
130 69.1 68.2 102.2





#29

Tetrahydrofuran

Concen: 280.74 ug/l

RT: 5.013 min Scan# 644





Delta R.T. 0.000 min

Lab File: VX022328.D

Acq: 02 Jun 2021 10:35

Tgt Ion: 42 Resp: 163722

Ion	Ratio	Lower	Upper
42	100		
72	45.1	41.7	62.5
71	41.2	38.4	57.6



#30

Chloroform

Concen: 52.47 ug/l

RT: 5.098 min Scan# 658




Delta R.T. -0.007 min


Lab File: VX022328.D

Acq: 02 Jun 2021 10:35

Tgt Ion: 83 Resp: 116309


Ion	Ratio	Lower	Upper
83	100		
85	64.3	52.0	78.0








#31
Cyclohexane
Concen: 56.72 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.013 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35


Instrument :
MSVOA_X
ClientSampled :





Tgt Ion: 56 Resp: 99132
Ion Ratio Lower Upper
56 100
69 31.2 25.3 37.9
84 83.0 71.4 107.0




#32
1,1,1-Trichloroethane
Concen: 52.65 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35



Tgt Ion: 97 Resp: 99735
Ion Ratio Lower Upper
97 100
99 62.5 51.5 77.3
61 45.5 34.2 51.4





#33

1,2-Dichloroethane-d4

Concen: 53.42 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX022328.D

Acq: 02 Jun 2021 10:35

Tgt Ion: 65 Resp: 74299

Ion Ratio Lower Upper





65 100

67 53.5 0.0 105.4

Instrument :

MSVOA_X

ClientSampled :



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX022328.D

Acq: 02 Jun 2021 10:35




Tgt Ion: 114 Resp: 175096


Ion Ratio Lower Upper

114 100

63 20.7 0.0 40.8

88 16.2 0.0 32.8









#35
Dibromofluoromethane
Concen: 49.77 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:113 Resp: 55262
Ion Ratio Lower Upper
113 100
111 104.9 81.9 122.9
192 16.5 16.7 25.1#



#36
1,1-Dichloropropene
Concen: 55.24 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.006 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 75 Resp: 89440
Ion Ratio Lower Upper
75 100
110 34.4 18.1 54.4
77 30.6 25.3 37.9





#37

Ethyl Acetate

Concen: 55.98 ug/l

RT: 4.720 min Scan# 596





Delta R.T. -0.001 min

Lab File: VX022328.D

Acq: 02 Jun 2021 10:35

Tgt Ion:	43	Resp:	111101
Ion	Ratio	Lower	Upper
43	100		
61	13.7	12.8	19.2
70	10.7	10.5	15.7

Instrument :
MSVOA_X
ClientSampled :



#38

Carbon Tetrachloride

Concen: 51.86 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.006 min

Lab File: VX022328.D

Acq: 02 Jun 2021 10:35

Tgt Ion:	117	Resp:	82145
Ion	Ratio	Lower	Upper
117	100		
119	88.9	76.2	114.2
121	29.9	24.8	37.2









#39
Methylcyclohexane
Concen: 56.03 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 102897
Ion Ratio Lower Upper
83 100
55 77.5 58.9 88.3
98 49.5 38.1 57.1



#40
Benzene
Concen: 54.63 ug/l
RT: 6.043 min Scan# 813
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 78 Resp: 272142
Ion Ratio Lower Upper
78 100
77 22.6 18.9 28.3









#41
Methacrylonitrile
Concen: 55.84 ug/l
RT: 4.922 min Scan# 629
Delta R.T. -0.006 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	41	Resp:	59288
Ion	Ratio	Lower	Upper
41	100		
39	53.3	42.5	63.7
67	73.6	65.0	97.6
52	29.3	25.3	37.9



#42
1,2-Dichloroethane
Concen: 54.58 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.007 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion:	62	Resp:	95250
Ion	Ratio	Lower	Upper
62	100		
98	9.4	0.0	20.2









#43
Isopropyl Acetate
Concen: 56.79 ug/l
RT: 6.348 min Scan# 863
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 184354
Ion Ratio Lower Upper
43 100
61 19.5 18.2 27.4
87 12.6 12.5 18.7



#44
Trichloroethene
Concen: 52.56 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion:130 Resp: 65064
Ion Ratio Lower Upper
130 100
95 107.1 0.0 198.0









#45
1,2-Dichloropropane
Concen: 56.55 ug/l
RT: 7.433 min Scan# 1041
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 63 Resp: 71864
Ion Ratio Lower Upper
63 100
65 30.9 25.7 38.5



#46
Dibromomethane
Concen: 55.09 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 93 Resp: 47499
Ion Ratio Lower Upper
93 100
95 80.5 66.8 100.2
174 79.7 83.1 124.7#









#47
Bromodichloromethane
Concen: 52.14 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 89947
Ion Ratio Lower Upper
83 100
85 66.9 51.4 77.0
127 7.7 6.8 10.2



#48
Methyl methacrylate
Concen: 57.35 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 41 Resp: 85797
Ion Ratio Lower Upper
41 100
69 83.2 76.0 114.0
39 59.5 44.6 67.0





#49

1,4-Dioxane

Concen: 1118.28 ug/l





RT: 7.659 min Scan# 1078

Delta R.T. -0.025 min

Lab File: VX022328.D

Acq: 02 Jun 2021 10:35

Tgt Ion:	88	Resp:	32575
Ion	Ratio	Lower	Upper
88	100		
43	33.4	23.2	34.8
58	72.1	55.0	82.4



#50

Toluene-d8

Concen: 51.18 ug/l




RT: 8.653 min Scan# 1241


Delta R.T. -0.000 min

Lab File: VX022328.D

Acq: 02 Jun 2021 10:35

Tgt Ion:	98	Resp:	218914
Ion	Ratio	Lower	Upper
98	100		
100	65.3	51.9	77.9









#51
4-Methyl-2-Pentanone
Concen: 291.68 ug/l
RT: 8.579 min Scan# 1229
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 567220
Ion Ratio Lower Upper
43 100
58 38.3 35.4 53.0



#52
Toluene
Concen: 53.77 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.006 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 92 Resp: 169623
Ion Ratio Lower Upper
92 100
91 170.3 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 54.02 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 75 Resp: 107578
Ion Ratio Lower Upper
75 100
77 32.9 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 55.26 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 75 Resp: 118316
Ion Ratio Lower Upper
75 100
77 29.3 25.7 38.5
39 45.7 35.1 52.7









#55
1,1,2-Trichloroethane
Concen: 53.68 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.006 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 97	Resp: 68744
Ion Ratio	Lower Upper
97 100	
83 90.4	69.3 103.9
85 57.8	44.3 66.5
99 63.2	50.3 75.5



#56
Ethyl methacrylate
Concen: 55.88 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 69	Resp: 115668
Ion Ratio	Lower Upper
69 100	
41 67.5	46.8 70.2
39 39.4	24.7 37.1#









#57
1,3-Dichloropropane
Concen: 55.39 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 121127
Ion Ratio Lower Upper
76 100
78 31.9 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 249.24 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.007 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 63 Resp: 279020
Ion Ratio Lower Upper
63 100
106 27.3 23.5 35.3





#59

2-Hexanone

Concen: 298.38 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX022328.D





Acq: 02 Jun 2021 10:35

Tgt Ion: 43 Resp: 436406

Ion Ratio Lower Upper

43 100

58 54.1 30.9 92.8



#60

Dibromochloromethane

Concen: 51.83 ug/l

RT: 9.524 min Scan# 1384

Delta R.T. -0.001 min

Lab File: VX022328.D




Acq: 02 Jun 2021 10:35


Tgt Ion:129 Resp: 64579

Ion Ratio Lower Upper

129 100

127 78.0 38.4 115.2









#61
1,2-Dibromoethane
Concen: 55.68 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:107 Resp: 71000
Ion Ratio Lower Upper
107 100
109 94.9 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 53.33 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 95 Resp: 84315
Ion Ratio Lower Upper
95 100
174 64.2 0.0 154.6
176 61.9 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:117 Resp: 161547
Ion Ratio Lower Upper
117 100
82 59.7 47.4 71.2
119 31.5 26.6 39.8



#64
Tetrachloroethene
Concen: 53.04 ug/l
RT: 9.274 min Scan# 1343
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion:164 Resp: 56254
Ion Ratio Lower Upper
164 100
166 124.4 103.4 155.2
129 111.0 70.8 106.2#
131 100.9 72.5 108.7









#65
Chlorobenzene
Concen: 51.91 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.007 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:112 Resp: 174575
Ion Ratio Lower Upper
112 100
114 31.8 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 52.03 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion:131 Resp: 60794
Ion Ratio Lower Upper
131 100
133 93.9 47.7 143.1
119 68.0 32.1 96.5









#67
Ethyl Benzene
Concen: 53.15 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 328827
Ion Ratio Lower Upper
91 100
106 30.5 24.6 37.0



#68
m/p-Xylenes
Concen: 105.54 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 106 Resp: 246759
Ion Ratio Lower Upper
106 100
91 205.5 162.8 244.2









#69
o-Xylene
Concen: 52.18 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:106 Resp: 118662
Ion Ratio Lower Upper
106 100
91 221.3 108.1 324.2



#70
Styrene
Concen: 54.65 ug/l
RT: 10.658 min Scan# 1570
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion:104 Resp: 203567
Ion Ratio Lower Upper
104 100
78 52.5 40.0 60.0
103 54.0 43.4 65.0









#71
Bromoform
Concen: 45.02 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:173 Resp: 40182
Ion Ratio Lower Upper
173 100
175 46.9 23.9 71.8
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion:152 Resp: 73514
Ion Ratio Lower Upper
152 100
115 96.0 41.1 123.3
150 177.1 0.0 349.0









#73
Isopropylbenzene
Concen: 50.86 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:105 Resp: 316455
Ion Ratio Lower Upper
105 100
120 25.7 13.2 39.6



#74
N-ethyl acetate
Concen: 57.05 ug/l
RT: 10.847 min Scan# 1601
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 43 Resp: 168483
Ion Ratio Lower Upper
43 100
70 40.7 39.0 58.6
55 22.2 21.5 32.3
61 23.8 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 53.77 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 110898
Ion Ratio Lower Upper
83 100
131 8.9 5.1 15.2
85 63.4 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 71.07 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 75 Resp: 128889
Ion Ratio Lower Upper
75 100
77 32.4 19.1 57.5





#77

Bromobenzene

Concen: 50.56 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. -0.000 min

Lab File: VX022328.D

Acq: 02 Jun 2021 10:35

Tgt Ion:156 Resp: 66704

Ion Ratio Lower Upper

156 100





77 196.3 82.8 248.6

158 97.3 49.1 147.4

Instrument :

MSVOA_X

ClientSampled :



#78

n-propylbenzene

Concen: 52.82 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX022328.D




Acq: 02 Jun 2021 10:35


Tgt Ion: 91 Resp: 384949

Ion Ratio Lower Upper

91 100

120 21.7 11.5 34.5









#79
2-Chlorotoluene
Concen: 50.54 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 223381
Ion Ratio Lower Upper
91 100
126 31.7 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 51.60 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.006 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 105 Resp: 266082
Ion Ratio Lower Upper
105 100
120 47.6 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 49.42 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 37828
Ion Ratio Lower Upper
75 100
53 98.8 74.5 111.7
89 43.6 37.9 56.9



#82
4-Chlorotoluene
Concen: 51.10 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 91 Resp: 260220
Ion Ratio Lower Upper
91 100
126 27.9 14.9 44.5









#83
tert-Butylbenzene
Concen: 50.09 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 247727
Ion Ratio Lower Upper
119 100
91 63.3 27.9 83.7
134 22.6 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 52.29 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion:105 Resp: 271433
Ion Ratio Lower Upper
105 100
120 43.4 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 52.36 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. -0.000 min
 Lab File: VX022328.D
 Acq: 02 Jun 2021 10:35




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion:105 Resp: 329084
 Ion Ratio Lower Upper
 105 100
 134 18.5 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 51.15 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX022328.D
 Acq: 02 Jun 2021 10:35

Tgt Ion:119 Resp: 271496
 Ion Ratio Lower Upper
 119 100
 134 25.9 12.9 38.7
 91 26.9 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 51.70 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.007 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 131513
Ion Ratio Lower Upper
146 100
111 43.7 20.2 60.5
148 61.6 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 49.68 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion:146 Resp: 130611
Ion Ratio Lower Upper
146 100
111 44.2 19.6 58.8
148 63.3 31.6 95.0









#89
n-Butylbenzene
Concen: 54.04 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 265328
Ion Ratio Lower Upper
91 100
92 54.7 27.4 82.0
134 23.6 13.2 39.5



#90
Hexachloroethane
Concen: 48.21 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. -0.001 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 117 Resp: 42290
Ion Ratio Lower Upper
117 100
201 55.0 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 50.49 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 123355
Ion Ratio Lower Upper
146 100
111 47.1 20.8 62.2
148 64.4 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 52.86 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion: 75 Resp: 23326
Ion Ratio Lower Upper
75 100
155 71.2 45.3 135.9
157 95.4 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 51.86 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:180 Resp: 76455
Ion Ratio Lower Upper
180 100
182 92.8 47.7 143.1
145 34.1 15.0 45.0



#94
Hexachlorobutadiene
Concen: 45.61 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.006 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion:225 Resp: 26245
Ion Ratio Lower Upper
225 100
223 63.7 31.1 93.2
227 66.2 31.6 95.0









#95
Naphthalene
Concen: 54.17 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:128 Resp: 311454
Ion Ratio Lower Upper
128 100
127 12.9 10.3 15.5
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 50.80 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX022328.D
Acq: 02 Jun 2021 10:35

Tgt Ion:180 Resp: 75498
Ion Ratio Lower Upper
180 100
182 95.0 47.6 142.9
145 37.8 16.1 48.3

