

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022330.D
 Acq On : 02 Jun 2021 11:50
 Operator : JC/MD
 Sample : VX0602WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0602WBS01

Quant Time: Jun 03 02:17:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	88301	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	182330	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	159945	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67590	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	74211	54.48	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.96%
35) Dibromofluoromethane	5.391	113	56393	48.77	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.54%
50) Toluene-d8	8.652	98	214895	48.25	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.50%
62) 4-Bromofluorobenzene	11.085	95	79413	48.24	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.48%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	23825	22.19	ug/l	100
3) Chloromethane	1.294	50	25935	22.40	ug/l	97
4) Vinyl Chloride	1.373	62	26991	22.03	ug/l	99
5) Bromomethane	1.599	94	18412	25.77	ug/l	92
6) Chloroethane	1.684	64	16227	22.24	ug/l	96
7) Trichlorofluoromethane	1.885	101	37501	21.29	ug/l	95
8) Diethyl Ether	2.135	74	16180	21.36	ug/l	90
9) 1,1,2-Trichlorotrifluo...	2.330	101	22820	21.81	ug/l	90
10) Methyl Iodide	2.452	142	24037	21.38	ug/l	95
11) Tert butyl alcohol	2.977	59	35401	113.63	ug/l	98
12) 1,1-Dichloroethene	2.318	96	23546	23.27	ug/l	94
13) Acrolein	2.239	56	22140	107.15	ug/l	100
14) Allyl chloride	2.666	41	42208	22.71	ug/l	97
15) Acrylonitrile	3.068	53	72811	118.71	ug/l	98
16) Acetone	2.385	43	67694	114.14	ug/l	95
17) Carbon Disulfide	2.513	76	55176	20.82	ug/l	99
18) Methyl Acetate	2.708	43	35453	22.48	ug/l	93
19) Methyl tert-butyl Ether	3.123	73	77036	22.22	ug/l	97
20) Methylene Chloride	2.794	84	28261	21.46	ug/l	87
21) trans-1,2-Dichloroethene	3.093	96	25771	23.41	ug/l	98
22) Diisopropyl ether	3.775	45	82049	22.92	ug/l	90
23) Vinyl Acetate	3.727	43	401356	116.44	ug/l	95
24) 1,1-Dichloroethane	3.617	63	48349	23.01	ug/l	97
25) 2-Butanone	4.568	43	115613	120.49	ug/l	98
26) 2,2-Dichloropropane	4.482	77	39747	21.31	ug/l	97
27) cis-1,2-Dichloroethene	4.501	96	30328	22.65	ug/l	99
28) Bromochloromethane	4.909	49	23620	23.17	ug/l	83
29) Tetrahydrofuran	5.019	42	66893	117.13	ug/l	92
30) Chloroform	5.104	83	48782	22.47	ug/l	89
31) Cyclohexane	5.476	56	40147	23.45	ug/l	97
32) 1,1,1-Trichloroethane	5.391	97	40522	21.85	ug/l	98
36) 1,1-Dichloropropene	5.702	75	35781	21.22	ug/l	99
37) Ethyl Acetate	4.726	43	44415	21.49	ug/l	94
38) Carbon Tetrachloride	5.684	117	32341	19.61	ug/l	98
39) Methylcyclohexane	7.384	83	40951	21.42	ug/l	96
40) Benzene	6.043	78	111967	21.58	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022330.D
 Acq On : 02 Jun 2021 11:50
 Operator : JC/MD
 Sample : VX0602WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0602WBS01

Quant Time: Jun 03 02:17:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.934	41	23214	21.00	ug/l	93
42) 1,2-Dichloroethane	6.098	62	39367	21.66	ug/l	100
43) Isopropyl Acetate	6.348	43	73383	21.71	ug/l	93
44) Trichloroethene	7.134	130	27771	21.54	ug/l	97
45) 1,2-Dichloropropane	7.439	63	28814	21.77	ug/l	96
46) Dibromomethane	7.592	93	18920	21.07	ug/l #	87
47) Bromodichloromethane	7.829	83	35405	19.71	ug/l	99
48) Methyl methacrylate	7.695	41	34018	21.84	ug/l	91
49) 1,4-Dioxane	7.665	88	13067	430.78	ug/l	92
51) 4-Methyl-2-Pentanone	8.579	43	230909	114.03	ug/l	91
52) Toluene	8.720	92	68341	20.80	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	41741	20.13	ug/l	98
54) cis-1,3-Dichloropropene	8.372	75	44897	20.14	ug/l	92
55) 1,1,2-Trichloroethane	9.159	97	28481	21.36	ug/l	98
56) Ethyl methacrylate	9.122	69	44958	20.86	ug/l #	86
57) 1,3-Dichloropropane	9.311	76	49602	21.78	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.244	63	112225	96.27	ug/l	97
59) 2-Hexanone	9.433	43	172565	113.31	ug/l	91
60) Dibromochloromethane	9.524	129	23273	17.94	ug/l	97
61) 1,2-Dibromoethane	9.616	107	28837	21.72	ug/l	95
64) Tetrachloroethene	9.274	164	22863	21.77	ug/l	89
65) Chlorobenzene	10.085	112	71491	21.47	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.164	131	23549	20.35	ug/l	97
67) Ethyl Benzene	10.195	91	131961	21.54	ug/l	100
68) m/p-Xylenes	10.305	106	96467	41.67	ug/l	95
69) o-Xylene	10.646	106	47712	21.19	ug/l	99
70) Styrene	10.658	104	78423	21.26	ug/l	99
71) Bromoform	10.805	173	14011	16.77	ug/l #	100
73) Isopropylbenzene	10.963	105	124615	21.78	ug/l	99
74) N-amyl acetate	10.847	43	61515	22.65	ug/l	93
75) 1,1,2,2-Tetrachloroethane	11.213	83	42722	22.53	ug/l	99
76) 1,2,3-Trichloropropane	11.243	75	49041	29.41	ug/l	89
77) Bromobenzene	11.201	156	26338	21.71	ug/l	87
78) n-propylbenzene	11.304	91	147446	22.01	ug/l	98
79) 2-Chlorotoluene	11.365	91	85379	21.01	ug/l	99
80) 1,3,5-Trimethylbenzene	11.457	105	105234	22.20	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.024	75	12969	18.43	ug/l	92
82) 4-Chlorotoluene	11.457	91	99657	21.28	ug/l	95
83) tert-Butylbenzene	11.719	119	95271	20.95	ug/l	92
84) 1,2,4-Trimethylbenzene	11.756	105	103665	21.72	ug/l	99
85) sec-Butylbenzene	11.896	105	124510	21.55	ug/l	98
86) p-Isopropyltoluene	12.012	119	101774	20.85	ug/l	96
87) 1,3-Dichlorobenzene	11.975	146	47793	20.44	ug/l	95
88) 1,4-Dichlorobenzene	12.042	146	49838	20.62	ug/l	96
89) n-Butylbenzene	12.335	91	94337	20.90	ug/l	97
90) Hexachloroethane	12.542	117	14589	18.09	ug/l	74
91) 1,2-Dichlorobenzene	12.335	146	46803	20.84	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.944	75	8067	19.88	ug/l	78
93) 1,2,4-Trichlorobenzene	13.591	180	26302	19.40	ug/l	98
94) Hexachlorobutadiene	13.725	225	9955	18.82	ug/l	95
95) Naphthalene	13.780	128	113668	21.50	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	28136	20.59	ug/l	96


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
Data File : VX022330.D
Acq On : 02 Jun 2021 11:50
Operator : JC/MD
Sample : VX0602WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01

Quant Time: Jun 03 02:17:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.562 min Scan# 734

Delta R.T. -0.000 min

Lab File: VX022330.D





Acq: 02 Jun 2021 11:50

Tgt Ion:168 Resp: 88301

Ion Ratio Lower Upper

168 100

99 73.1 45.8 68.6#



#2

Dichlorodifluoromethane

Concen: 22.19 ug/l

RT: 1.166 min Scan# 13

Delta R.T. -0.001 min

Lab File: VX022330.D




Acq: 02 Jun 2021 11:50


Tgt Ion: 85 Resp: 23825

Ion Ratio Lower Upper

85 100

87 32.7 16.4 49.2









#3
Chloromethane
Concen: 22.40 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion: 50 Resp: 25935
Ion Ratio Lower Upper
50 100
52 31.5 26.4 39.6



#4
Vinyl Chloride
Concen: 22.03 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 62 Resp: 26991
Ion Ratio Lower Upper
62 100
64 31.5 25.7 38.5





#5

Bromomethane

Concen: 25.77 ug/l

RT: 1.599 min Scan# 84

Delta R.T. -0.013 min

Lab File: VX022330.D





Acq: 02 Jun 2021 11:50

Tgt Ion: 94 Resp: 18412

Ion Ratio Lower Upper

94 100

96 88.0 76.4 114.6



#6

Chloroethane

Concen: 22.24 ug/l

RT: 1.684 min Scan# 98

Delta R.T. -0.001 min

Lab File: VX022330.D




Acq: 02 Jun 2021 11:50


Tgt Ion: 64 Resp: 16227

Ion Ratio Lower Upper

64 100

66 33.8 25.4 38.0









#7
 Trichlorofluoromethane
 Concen: 21.29 ug/l
 RT: 1.885 min Scan# 131
 Delta R.T. -0.001 min
 Lab File: VX022330.D
 Acq: 02 Jun 2021 11:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0602WBS01

Tgt Ion:101 Resp: 37501
 Ion Ratio Lower Upper
 101 100
 103 70.5 53.2 79.8



#8
 Diethyl Ether
 Concen: 21.36 ug/l
 RT: 2.135 min Scan# 172
 Delta R.T. -0.001 min
 Lab File: VX022330.D
 Acq: 02 Jun 2021 11:50

Tgt Ion: 74 Resp: 16180
 Ion Ratio Lower Upper
 74 100
 45 98.8 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.81 ug/l
RT: 2.330 min Scan# 204
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion:	101	Resp:	22820
Ion	Ratio	Lower	Upper
101	100		
85	46.4	34.9	52.3
151	65.6	61.4	92.2



#10
Methyl Iodide
Concen: 21.38 ug/l
RT: 2.452 min Scan# 224
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion:	142	Resp:	24037
Ion	Ratio	Lower	Upper
142	100		
127	46.0	33.3	49.9
141	14.1	11.4	17.0









#11
Tert butyl alcohol
Concen: 113.63 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.012 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion: 59 Resp: 35401
Ion Ratio Lower Upper
59 100
57 10.6 7.8 11.8



#12
1,1-Dichloroethene
Concen: 23.27 ug/l
RT: 2.318 min Scan# 202
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 96 Resp: 23546
Ion Ratio Lower Upper
96 100
61 163.4 124.6 187.0
98 58.8 50.6 76.0









#13
Acrolein
Concen: 107.15 ug/l
RT: 2.239 min Scan# 189
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01

Tgt Ion: 56 Resp: 22140
Ion Ratio Lower Upper
56 100
55 70.5 56.4 84.6



#14
Allyl chloride
Concen: 22.71 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 41 Resp: 42208
Ion Ratio Lower Upper
41 100
39 65.3 51.6 77.4
76 34.1 30.5 45.7









#15
Acrylonitrile
Concen: 118.71 ug/l
RT: 3.068 min Scan# 325
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion:	53	Resp:	72811
Ion	Ratio	Lower	Upper
53	100		
52	84.0	65.3	97.9
51	35.3	28.2	42.2



#16
Acetone
Concen: 114.14 ug/l
RT: 2.385 min Scan# 213
Delta R.T. -0.007 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion:	43	Resp:	67694
Ion	Ratio	Lower	Upper
43	100		
58	32.2	27.8	41.8









#17
Carbon Disulfide
Concen: 20.82 ug/l
RT: 2.513 min Scan# 234
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion: 76 Resp: 55176
Ion Ratio Lower Upper
76 100
78 8.8 7.2 10.8



#18
Methyl Acetate
Concen: 22.48 ug/l
RT: 2.708 min Scan# 266
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 43 Resp: 35453
Ion Ratio Lower Upper
43 100
74 24.0 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 22.22 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.006 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion: 73 Resp: 77036
Ion Ratio Lower Upper
73 100
57 22.7 17.1 25.7



#20
Methylene Chloride
Concen: 21.46 ug/l
RT: 2.794 min Scan# 280
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 84 Resp: 28261
Ion Ratio Lower Upper
84 100
49 130.3 88.6 132.8
51 38.4 27.0 40.6
86 68.7 50.3 75.5








#21
trans-1,2-Dichloroethene
Concen: 23.41 ug/l
RT: 3.093 min Scan# 329
Delta R.T. -0.006 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion: 96 Resp: 25771
Ion Ratio Lower Upper
96 100
61 133.2 107.7 161.5
98 66.4 50.4 75.6



#22
Diisopropyl ether
Concen: 22.92 ug/l
RT: 3.775 min Scan# 441
Delta R.T. 0.005 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 45 Resp: 82049
Ion Ratio Lower Upper
45 100
43 61.3 58.0 87.0
87 28.3 25.4 38.0
59 10.9 10.2 15.2









#23
 Vinyl Acetate
 Concen: 116.44 ug/l
 RT: 3.727 min Scan# 433
 Delta R.T. -0.000 min
 Lab File: VX022330.D
 Acq: 02 Jun 2021 11:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0602WBS01


Tgt Ion: 43 Resp: 401356
 Ion Ratio Lower Upper
 43 100
 86 10.5 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 23.01 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. -0.000 min
 Lab File: VX022330.D
 Acq: 02 Jun 2021 11:50

Tgt Ion: 63 Resp: 48349
 Ion Ratio Lower Upper
 63 100
 98 6.6 3.8 11.2
 100 3.5 2.5 7.5









#25
2-Butanone
Concen: 120.49 ug/l
RT: 4.568 min Scan# 571
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampled :
VX0602WBS01

Tgt Ion: 43 Resp: 115613
Ion Ratio Lower Upper
43 100
72 28.7 23.7 35.5



#26
2,2-Dichloropropane
Concen: 21.31 ug/l
RT: 4.482 min Scan# 557
Delta R.T. 0.005 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 77 Resp: 39747
Ion Ratio Lower Upper
77 100
97 22.0 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 22.65 ug/l

RT: 4.501 min Scan# 560

Delta R.T. 0.006 min

Lab File: VX022330.D

Acq: 02 Jun 2021 11:50

Tgt Ion: 96 Resp: 30328





Ion	Ratio	Lower	Upper
96	100		
61	144.6	0.0	285.6
98	64.9	0.0	129.6

Instrument :

MSVOA_X

ClientSampleId :

VX0602WBS01



#28

Bromochloromethane

Concen: 23.17 ug/l

RT: 4.909 min Scan# 627




Delta R.T. -0.001 min


Lab File: VX022330.D

Acq: 02 Jun 2021 11:50

Tgt Ion: 49 Resp: 23620

Ion	Ratio	Lower	Upper
49	100		
129	1.5	0.0	4.0
130	69.3	68.2	102.2





#29

Tetrahydrofuran

Concen: 117.13 ug/l

RT: 5.019 min Scan# 645

Delta R.T. 0.006 min

Lab File: VX022330.D

Acq: 02 Jun 2021 11:50





Tgt Ion: 42 Resp: 66893

Ion Ratio Lower Upper

42 100

72 45.9 41.7 62.5

71 43.4 38.4 57.6



#30

Chloroform

Concen: 22.47 ug/l

RT: 5.104 min Scan# 659

Delta R.T. -0.001 min

Lab File: VX022330.D




Acq: 02 Jun 2021 11:50


Tgt Ion: 83 Resp: 48782

Ion Ratio Lower Upper

83 100

85 55.9 52.0 78.0









#31
Cyclohexane
Concen: 23.45 ug/l
RT: 5.476 min Scan# 720
Delta R.T. -0.007 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01

Tgt Ion: 56 Resp: 40147




Ion	Ratio	Lower	Upper
56	100		
69	30.8	25.3	37.9
84	92.7	71.4	107.0




#32
1,1,1-Trichloroethane
Concen: 21.85 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 97 Resp: 40522

Ion	Ratio	Lower	Upper
97	100		
99	63.9	51.5	77.3
61	46.1	34.2	51.4





#33

1,2-Dichloroethane-d4

Concen: 54.48 ug/l

RT: 5.964 min Scan# 800

Delta R.T. -0.000 min

Lab File: VX022330.D





Acq: 02 Jun 2021 11:50

Tgt Ion: 65 Resp: 74211

Ion Ratio Lower Upper

65 100

67 52.4 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.769 min Scan# 932

Delta R.T. -0.000 min

Lab File: VX022330.D

Acq: 02 Jun 2021 11:50




Tgt Ion: 114 Resp: 182330

Ion Ratio Lower Upper

114 100

63 20.4 0.0 40.8

88 15.6 0.0 32.8









#35
Dibromofluoromethane
Concen: 48.77 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion:113 Resp: 56393
Ion Ratio Lower Upper
113 100
111 102.4 81.9 122.9
192 15.4 16.7 25.1#



#36
1,1-Dichloropropene
Concen: 21.22 ug/l
RT: 5.702 min Scan# 757
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 75 Resp: 35781
Ion Ratio Lower Upper
75 100
110 35.1 18.1 54.4
77 31.2 25.3 37.9





#37

Ethyl Acetate

Concen: 21.49 ug/l

RT: 4.726 min Scan# 597





Delta R.T. 0.005 min

Lab File: VX022330.D

Acq: 02 Jun 2021 11:50

Tgt Ion:	43	Resp:	44415
Ion Ratio	Lower	Upper	
43	100		
61	13.0	12.8	19.2
70	11.0	10.5	15.7

Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01



#38

Carbon Tetrachloride

Concen: 19.61 ug/l




RT: 5.684 min Scan# 754

Delta R.T. -0.006 min

Lab File: VX022330.D

Acq: 02 Jun 2021 11:50

Tgt Ion:	117	Resp:	32341
Ion Ratio	Lower	Upper	
117	100		
119	93.4	76.2	114.2
121	29.0	24.8	37.2





#39

Methylcyclohexane

Concen: 21.42 ug/l

RT: 7.384 min Scan# 1033





Delta R.T. -0.001 min

Lab File: VX022330.D

Acq: 02 Jun 2021 11:50

Tgt Ion:	83	Resp:	40951
Ion	Ratio	Lower	Upper
83	100		
55	79.0	58.9	88.3
98	48.2	38.1	57.1

Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01



#40

Benzene

Concen: 21.58 ug/l




RT: 6.043 min Scan# 813

Delta R.T. -0.001 min

Lab File: VX022330.D

Acq: 02 Jun 2021 11:50

Tgt Ion:	78	Resp:	111967
Ion	Ratio	Lower	Upper
78	100		
77	22.7	18.9	28.3









#41
Methacrylonitrile
Concen: 21.00 ug/l
RT: 4.934 min Scan# 631
Delta R.T. 0.006 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion:	41	Resp:	23214
Ion	Ratio	Lower	Upper
41	100		
39	58.9	42.5	63.7
67	74.1	65.0	97.6
52	30.7	25.3	37.9



#42
1,2-Dichloroethane
Concen: 21.66 ug/l
RT: 6.098 min Scan# 822
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion:	62	Resp:	39367
Ion	Ratio	Lower	Upper
62	100		
98	9.9	0.0	20.2









#43
Isopropyl Acetate
Concen: 21.71 ug/l
RT: 6.348 min Scan# 863
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01

Tgt Ion: 43	Resp: 73383
Ion Ratio	Lower Upper
43	100
61	19.4 18.2 27.4
87	12.7 12.5 18.7



#44
Trichloroethene
Concen: 21.54 ug/l
RT: 7.134 min Scan# 992
Delta R.T. 0.005 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 130	Resp: 27771
Ion Ratio	Lower Upper
130	100
95	101.6 0.0 198.0









#45

1,2-Dichloropropane
Concen: 21.77 ug/l
RT: 7.439 min Scan# 1042
Delta R.T. 0.005 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Instrument :
MSVOA_X
ClientSampled :
VX0602WBS01




Tgt Ion: 63 Resp: 28814
Ion Ratio Lower Upper
63 100
65 30.0 25.7 38.5



#46

Dibromomethane
Concen: 21.07 ug/l
RT: 7.592 min Scan# 1067
Delta R.T. 0.006 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 93 Resp: 18920
Ion Ratio Lower Upper
93 100
95 82.7 66.8 100.2
174 80.9 83.1 124.7#









#47
Bromodichloromethane
Concen: 19.71 ug/l
RT: 7.829 min Scan# 1106
Delta R.T. 0.005 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01

Tgt Ion: 83	Resp: 35405
Ion Ratio	Lower Upper
83 100	
85 63.2	51.4 77.0
127 7.9	6.8 10.2



#48
Methyl methacrylate
Concen: 21.84 ug/l
RT: 7.695 min Scan# 1084
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 41	Resp: 34018
Ion Ratio	Lower Upper
41 100	
69 84.1	76.0 114.0
39 60.0	44.6 67.0









#49
1,4-Dioxane
Concen: 430.78 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. -0.019 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion: 88 Resp: 13067
Ion Ratio Lower Upper
88 100
43 34.4 23.2 34.8
58 74.1 55.0 82.4



#50
Toluene-d8
Concen: 48.25 ug/l
RT: 8.652 min Scan# 1241
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 98 Resp: 214895
Ion Ratio Lower Upper
98 100
100 66.3 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 114.03 ug/l
RT: 8.579 min Scan# 1229
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion: 43 Resp: 230909
Ion Ratio Lower Upper
43 100
58 38.4 35.4 53.0



#52
Toluene
Concen: 20.80 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.006 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 92 Resp: 68341
Ion Ratio Lower Upper
92 100
91 174.3 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 20.13 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion: 75 Resp: 41741
Ion Ratio Lower Upper
75 100
77 30.2 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 20.14 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 75 Resp: 44897
Ion Ratio Lower Upper
75 100
77 31.1 25.7 38.5
39 52.4 35.1 52.7









#55
1,1,2-Trichloroethane
Concen: 21.36 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion:	97	Resp:	28481
Ion	Ratio	Lower	Upper
97	100		
83	83.6	69.3	103.9
85	57.4	44.3	66.5
99	63.2	50.3	75.5



#56
Ethyl methacrylate
Concen: 20.86 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion:	69	Resp:	44958
Ion	Ratio	Lower	Upper
69	100		
41	67.1	46.8	70.2
39	41.9	24.7	37.1









#57

1,3-Dichloropropane
Concen: 21.78 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01




Tgt Ion: 76 Resp: 49602
Ion Ratio Lower Upper
76 100
78 31.3 26.2 39.2




#58

2-Chloroethyl Vinyl ether
Concen: 96.27 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.007 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 63 Resp: 112225
Ion Ratio Lower Upper
63 100
106 27.6 23.5 35.3









#59
2-Hexanone
Concen: 113.31 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion: 43 Resp: 172565
Ion Ratio Lower Upper
43 100
58 54.9 30.9 92.8



#60
Dibromochloromethane
Concen: 17.94 ug/l
RT: 9.524 min Scan# 1384
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 129 Resp: 23273
Ion Ratio Lower Upper
129 100
127 79.1 38.4 115.2









#61
1,2-Dibromoethane
Concen: 21.72 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion:107 Resp: 28837
Ion Ratio Lower Upper
107 100
109 90.1 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 48.24 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 95 Resp: 79413
Ion Ratio Lower Upper
95 100
174 63.3 0.0 154.6
176 60.9 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion:117 Resp: 159945
Ion Ratio Lower Upper
117 100
82 61.4 47.4 71.2
119 31.5 26.6 39.8



#64
Tetrachloroethene
Concen: 21.77 ug/l
RT: 9.274 min Scan# 1343
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion:164 Resp: 22863
Ion Ratio Lower Upper
164 100
166 113.9 103.4 155.2
129 103.0 70.8 106.2
131 92.4 72.5 108.7









#65
Chlorobenzene
Concen: 21.47 ug/l
RT: 10.085 min Scan# 1476
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampled :
VX0602WBS01


Tgt Ion:112 Resp: 71491
Ion Ratio Lower Upper
112 100
114 31.9 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 20.35 ug/l
RT: 10.164 min Scan# 1489
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion:131 Resp: 23549
Ion Ratio Lower Upper
131 100
133 94.8 47.7 143.1
119 69.2 32.1 96.5









#67
Ethyl Benzene
Concen: 21.54 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion: 91 Resp: 131961
Ion Ratio Lower Upper
91 100
106 30.6 24.6 37.0



#68
m/p-Xylenes
Concen: 41.67 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion:106 Resp: 96467
Ion Ratio Lower Upper
106 100
91 210.7 162.8 244.2









#69
o-Xylene
Concen: 21.19 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01

Tgt Ion:106 Resp: 47712
Ion Ratio Lower Upper
106 100
91 217.5 108.1 324.2



#70
Styrene
Concen: 21.26 ug/l
RT: 10.658 min Scan# 1570
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion:104 Resp: 78423
Ion Ratio Lower Upper
104 100
78 51.8 40.0 60.0
103 54.5 43.4 65.0





#71

Bromoform

Concen: 16.77 ug/l

RT: 10.805 min Scan# 1594





Delta R.T. -0.000 min

Lab File: VX022330.D

Acq: 02 Jun 2021 11:50

Tgt Ion:	173	Resp:	14011
Ion Ratio	Lower	Upper	
173	100		
175	47.8	23.9	71.8
254	0.0	0.1	0.1#

Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. -0.000 min

Lab File: VX022330.D

Acq: 02 Jun 2021 11:50

Tgt Ion:	152	Resp:	67590
Ion Ratio	Lower	Upper	
152	100		
115	79.2	41.1	123.3
150	169.0	0.0	349.0









#73
Isopropylbenzene
Concen: 21.78 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion:105 Resp: 124615
Ion Ratio Lower Upper
105 100
120 25.9 13.2 39.6



#74
N-ethyl acetate
Concen: 22.65 ug/l
RT: 10.847 min Scan# 1601
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 43 Resp: 61515
Ion Ratio Lower Upper
43 100
70 43.6 39.0 58.6
55 24.2 21.5 32.3
61 24.3 22.1 33.1









#75

1,1,2,2-Tetrachloroethane
Concen: 22.53 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01




Tgt Ion: 83 Resp: 42722
Ion Ratio Lower Upper
83 100
131 9.0 5.1 15.2
85 63.5 31.8 95.4




#76

1,2,3-Trichloropropane
Concen: 29.41 ug/l
RT: 11.243 min Scan# 1666
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 75 Resp: 49041
Ion Ratio Lower Upper
75 100
77 31.4 19.1 57.5









#77
Bromobenzene
Concen: 21.71 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion:156 Resp: 26338
Ion Ratio Lower Upper
156 100
77 192.0 82.8 248.6
158 96.8 49.1 147.4



#78
n-propylbenzene
Concen: 22.01 ug/l
RT: 11.304 min Scan# 1676
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 91 Resp: 147446
Ion Ratio Lower Upper
91 100
120 21.9 11.5 34.5









#79
2-Chlorotoluene
Concen: 21.01 ug/l
RT: 11.365 min Scan# 1686
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01

Tgt Ion: 91 Resp: 85379
Ion Ratio Lower Upper
91 100
126 32.4 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 22.20 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 105 Resp: 105234
Ion Ratio Lower Upper
105 100
120 46.4 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 18.43 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion: 75 Resp: 12969
Ion Ratio Lower Upper
75 100
53 101.4 74.5 111.7
89 43.7 37.9 56.9



#82
4-Chlorotoluene
Concen: 21.28 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 91 Resp: 99657
Ion Ratio Lower Upper
91 100
126 27.2 14.9 44.5









#83
tert-Butylbenzene
Concen: 20.95 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion:119	Resp:	95271
Ion	Ratio	Lower Upper
119	100	
91	63.3	27.9 83.7
134	23.4	11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 21.72 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion:105	Resp:	103665
Ion	Ratio	Lower Upper
105	100	
120	44.4	22.4 67.3









#85
 sec-Butylbenzene
 Concen: 21.55 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. -0.000 min
 Lab File: VX022330.D
 Acq: 02 Jun 2021 11:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0602WBS01


Tgt Ion:105 Resp: 124510
 Ion Ratio Lower Upper
 105 100
 134 19.0 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 20.85 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX022330.D
 Acq: 02 Jun 2021 11:50

Tgt Ion:119 Resp: 101774
 Ion Ratio Lower Upper
 119 100
 134 25.2 12.9 38.7
 91 26.4 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 20.44 ug/l
RT: 11.975 min Scan# 1786
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion:146 Resp: 47793
Ion Ratio Lower Upper
146 100
111 47.5 20.2 60.5
148 64.1 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 20.62 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion:146 Resp: 49838
Ion Ratio Lower Upper
146 100
111 44.3 19.6 58.8
148 64.4 31.6 95.0









#89
n-Butylbenzene
Concen: 20.90 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion: 91 Resp: 94337
Ion Ratio Lower Upper
91 100
92 52.8 27.4 82.0
134 23.6 13.2 39.5



#90
Hexachloroethane
Concen: 18.09 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 117 Resp: 14589
Ion Ratio Lower Upper
117 100
201 57.6 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 20.84 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.006 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01


Tgt Ion:146 Resp: 46803
Ion Ratio Lower Upper
146 100
111 47.6 20.8 62.2
148 64.0 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.88 ug/l
RT: 12.944 min Scan# 1945
Delta R.T. -0.001 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion: 75 Resp: 8067
Ion Ratio Lower Upper
75 100
155 72.8 45.3 135.9
157 89.3 57.7 173.1









#93
 1,2,4-Trichlorobenzene
 Concen: 19.40 ug/l
 RT: 13.591 min Scan# 2051
 Delta R.T. -0.000 min
 Lab File: VX022330.D
 Acq: 02 Jun 2021 11:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0602WBS01


Tgt Ion:180 Resp: 26302
 Ion Ratio Lower Upper
 180 100
 182 95.8 47.7 143.1
 145 35.0 15.0 45.0



#94
 Hexachlorobutadiene
 Concen: 18.82 ug/l
 RT: 13.725 min Scan# 2073
 Delta R.T. -0.006 min
 Lab File: VX022330.D
 Acq: 02 Jun 2021 11:50

Tgt Ion:225 Resp: 9955
 Ion Ratio Lower Upper
 225 100
 223 59.5 31.1 93.2
 227 58.0 31.6 95.0









#95
Naphthalene
Concen: 21.50 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0602WBS01

Tgt Ion:128 Resp: 113668
Ion Ratio Lower Upper
128 100
127 12.8 10.3 15.5
129 10.8 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 20.59 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX022330.D
Acq: 02 Jun 2021 11:50

Tgt Ion:180 Resp: 28136
Ion Ratio Lower Upper
180 100
182 91.6 47.6 142.9
145 34.3 16.1 48.3

