

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022332.D
 Acq On : 02 Jun 2021 12:43
 Operator : JC/MD
 Sample : M2561-02DL2 800X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 36BR-20210526DL2

Quant Time: Jun 03 02:19:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	155465	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	137052	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51355	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	62196	55.311	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 110.620%			
35) Dibromofluoromethane	5.391	113	47297	47.974	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.940%			
50) Toluene-d8	8.653	98	190079	50.051	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.100%			
62) 4-Bromofluorobenzene	11.085	95	66002	47.019	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.040%			
Target Compounds						
4) Vinyl Chloride	1.374	62	8066	7.974	ug/l	97
12) 1,1-Dichloroethene	2.313	96	283	0.339	ug/l	91
27) cis-1,2-Dichloroethene	4.495	96	54134	48.983	ug/l	92
44) Trichloroethene	7.129	130	31062	28.260	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
Data File : VX022332.D
Acq On : 02 Jun 2021 12:43
Operator : JC/MD
Sample : M2561-02DL2 800X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
36BR-20210526DL2

Quant Time: Jun 03 02:19:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

