

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022336.D
 Acq On : 02 Jun 2021 14:17
 Operator : JC/MD
 Sample : M2572-05DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20BR-20210528DL

Quant Time: Jun 03 02:20:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64851	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	142717	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123184	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44770	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	57225	57.204	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.400%
35) Dibromofluoromethane	5.397	113	42840	47.335	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.660%
50) Toluene-d8	8.653	98	173689	49.820	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.640%
62) 4-Bromofluorobenzene	11.085	95	58522	45.415	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.820%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	314	0.398	ug/l	72
4) Vinyl Chloride	1.374	62	8314	9.239	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	54380	55.310	ug/l	91
44) Trichloroethene	7.129	130	13580	13.459	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
Data File : VX022336.D
Acq On : 02 Jun 2021 14:17
Operator : JC/MD
Sample : M2572-05DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20BR-20210528DL

Quant Time: Jun 03 02:20:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

