

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022337.D
 Acq On : 02 Jun 2021 14:41
 Operator : JC/MD
 Sample : M2572-10DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 29BR-20210528DL

Quant Time: Jun 03 02:20:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	62384	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	131703	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	117059	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45932	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53776	55.882	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.760%
35) Dibromofluoromethane	5.397	113	39476	47.265	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.540%
50) Toluene-d8	8.653	98	163265	50.746	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.500%
62) 4-Bromofluorobenzene	11.085	95	56567	47.569	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.140%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	294	0.388	ug/l	76
9) 1,1,2-Trichlorotrifluo...	2.325	101	288	0.390	ug/l #	77
12) 1,1-Dichloroethene	2.319	96	228	0.319	ug/l	83
27) cis-1,2-Dichloroethene	4.489	96	3861	4.082	ug/l	89
44) Trichloroethene	7.135	130	37834	40.632	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
Data File : VX022337.D
Acq On : 02 Jun 2021 14:41
Operator : JC/MD
Sample : M2572-10DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
29BR-20210528DL

Quant Time: Jun 03 02:20:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

