

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022338.D
 Acq On : 02 Jun 2021 15:04
 Operator : JC/MD
 Sample : M2572-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB03-20210527

Quant Time: Jun 03 02:20:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	63344	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	131539	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	117559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46639	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54440	55.714	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.420%
35) Dibromofluoromethane	5.397	113	39854	47.778	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.560%
50) Toluene-d8	8.653	98	164750	51.272	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.540%
62) 4-Bromofluorobenzene	11.085	95	56747	47.779	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.560%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.965	59	2469	11.048	ug/l #	91
16) Acetone	2.386	43	4981	11.707	ug/l #	86
25) 2-Butanone	4.574	43	4606	6.692	ug/l	95
30) Chloroform	5.117	83	458	0.294	ug/l #	50
44) Trichloroethene	7.135	130	287	0.309	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
Data File : VX022338.D
Acq On : 02 Jun 2021 15:04
Operator : JC/MD
Sample : M2572-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB03-20210527

Quant Time: Jun 03 02:20:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

