

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022339.D
 Acq On : 02 Jun 2021 15:28
 Operator : JC/MD
 Sample : M2572-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB02-20210527

Quant Time: Jun 03 02:21:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61978	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	135513	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	120181	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46682	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	54264	56.758	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.520%
35) Dibromofluoromethane	5.391	113	40208	46.788	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.580%
50) Toluene-d8	8.653	98	165257	49.921	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.840%
62) 4-Bromofluorobenzene	11.085	95	57035	46.614	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.220%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	2715	6.522	ug/l	99
20) Methylene Chloride	2.800	84	1120	1.212	ug/l	# 95
52) Toluene	8.726	92	1517	0.621	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
Data File : VX022339.D
Acq On : 02 Jun 2021 15:28
Operator : JC/MD
Sample : M2572-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB02-20210527

Quant Time: Jun 03 02:21:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

