

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022341.D
 Acq On : 02 Jun 2021 16:15
 Operator : JC/MD
 Sample : M2579-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33S-20210601

Quant Time: Jun 03 02:21:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61347	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	129273	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44316	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53558	56.596	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.200%
35) Dibromofluoromethane	5.397	113	38160	46.549	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.100%
50) Toluene-d8	8.653	98	159507	50.510	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.020%
62) 4-Bromofluorobenzene	11.085	95	54098	46.347	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.700%
Target Compounds						
16) Acetone	2.386	43	2661	6.458	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
Data File : VX022341.D
Acq On : 02 Jun 2021 16:15
Operator : JC/MD
Sample : M2579-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33S-20210601

Quant Time: Jun 03 02:21:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

