

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022343.D
 Acq On : 02 Jun 2021 17:02
 Operator : JC/MD
 Sample : M2572-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 91BR-20210528

Quant Time: Jun 03 02:22:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	62559	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	132445	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	53834	55.786	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.580%
35) Dibromofluoromethane	5.397	113	39935	47.547	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.100%
50) Toluene-d8	8.653	98	161856	50.027	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.060%
62) 4-Bromofluorobenzene	11.085	95	55683	46.563	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.120%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.173	85	5463	7.183	ug/l	99
4) Vinyl Chloride	1.374	62	196510	226.374	ug/l	100
12) 1,1-Dichloroethene	2.325	96	1467	2.046	ug/l #	69
16) Acetone	2.380	43	2107	5.014	ug/l #	88
21) trans-1,2-Dichloroethene	3.099	96	4319	5.538	ug/l	92
27) cis-1,2-Dichloroethene	4.495	96	657055	692.775	ug/l	91
44) Trichloroethene	7.129	130	8662	9.250	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
Data File : VX022343.D
Acq On : 02 Jun 2021 17:02
Operator : JC/MD
Sample : M2572-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
91BR-20210528

Quant Time: Jun 03 02:22:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

