

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022345.D
 Acq On : 02 Jun 2021 17:49
 Operator : JC/MD
 Sample : M2572-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 08BR-20210528

Quant Time: Jun 03 02:22:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59156	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	127332	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	113531	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42040	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51982	56.965	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.940%
35) Dibromofluoromethane	5.397	113	37800	46.812	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.620%
50) Toluene-d8	8.653	98	156005	50.154	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.300%
62) 4-Bromofluorobenzene	11.085	95	54123	47.076	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.160%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	7207	10.283	ug/l	91
16) Acetone	2.386	43	2246	5.653	ug/l	# 81
24) 1,1-Dichloroethane	3.617	63	1440	1.023	ug/l	# 88
27) cis-1,2-Dichloroethene	4.501	96	14727	16.421	ug/l	89
44) Trichloroethene	7.135	130	268915	298.713	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
Data File : VX022345.D
Acq On : 02 Jun 2021 17:49
Operator : JC/MD
Sample : M2572-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
08BR-20210528

Quant Time: Jun 03 02:22:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

