

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022328.D
 Acq On : 02 Jun 2021 10:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 03 02:17:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.608	0.632	-3.9	97	0.00
3 P	Chloromethane	0.656	0.697	-6.2	100	0.00
4 C	Vinyl Chloride	0.694	0.713	-2.7#	96	0.00
5 T	Bromomethane	0.404	0.404	0.0	102	-0.01
6 T	Chloroethane	0.413	0.430	-4.1	96	0.00
7 T	Trichlorofluoromethane	0.998	1.037	-3.9	99	0.00
8 T	Diethyl Ether	0.429	0.420	2.1	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.592	0.627	-5.9	102	0.00
10 T	Methyl Iodide	0.637	0.702	-10.2	102	0.00
11 T	Tert butyl alcohol	0.176	0.185	-5.1	105	-0.01
12 CM	1,1-Dichloroethene	0.573	0.613	-7.0#	102	0.00
13 T	Acrolein	0.117	0.124	-6.0	111	0.00
14 T	Allyl chloride	1.053	1.168	-10.9	106	0.00
15 T	Acrylonitrile	0.347	0.393	-13.3	107	0.00
16 T	Acetone	0.336	0.401	-19.3	121	0.00
17 T	Carbon Disulfide	1.501	1.605	-6.9	108	0.00
18 T	Methyl Acetate	0.893	0.980	-9.7	109	0.00
19 T	Methyl tert-butyl Ether	1.963	2.120	-8.0	103	0.00
20 T	Methylene Chloride	0.746	0.743	0.4	104	0.00
21 T	trans-1,2-Dichloroethene	0.623	0.693	-11.2	104	0.00
22 T	Diisopropyl ether	2.027	2.269	-11.9	108	0.00
23 T	Vinyl Acetate	1.952	2.254	-15.5	109	0.00
24 P	1,1-Dichloroethane	1.190	1.305	-9.7	105	0.00
25 T	2-Butanone	0.543	0.643	-18.4	113	0.00
26 T	2,2-Dichloropropane	1.056	1.095	-3.7	100	0.00
27 T	cis-1,2-Dichloroethene	0.758	0.816	-7.7	101	0.00
28 T	Bromochloromethane	0.577	0.627	-8.7	108	0.00
29 T	Tetrahydrofuran	0.323	0.363	-12.4	109	0.00
30 C	Chloroform	1.229	1.290	-5.0#	101	0.00
31 T	Cyclohexane	0.969	1.099	-13.4	111	-0.01
32 T	1,1,1-Trichloroethane	1.050	1.106	-5.3	101	0.00
33 S	1,2-Dichloroethane-d4	0.771	0.824	-6.9	114	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.317	0.316	0.3	105	0.00
36 T	1,1-Dichloropropene	0.462	0.511	-10.6	106	0.00
37 T	Ethyl Acetate	0.567	0.635	-12.0	112	0.00
38 T	Carbon Tetrachloride	0.452	0.469	-3.8	95	0.00
39 T	Methylcyclohexane	0.524	0.588	-12.2	110	0.00
40 TM	Benzene	1.423	1.554	-9.2	105	0.00
41 T	Methacrylonitrile	0.303	0.339	-11.9	109	0.00
42 TM	1,2-Dichloroethane	0.498	0.544	-9.2	105	0.00
43 T	Isopropyl Acetate	0.927	1.053	-13.6	111	0.00
44 TM	Trichloroethene	0.354	0.372	-5.1	101	0.00
45 C	1,2-Dichloropropane	0.363	0.410	-12.9#	104	0.00
46 T	Dibromomethane	0.246	0.271	-10.2	102	0.00
47 T	Bromodichloromethane	0.493	0.514	-4.3	96	0.00
48 T	Methyl methacrylate	0.427	0.490	-14.8	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
 Data File : VX022328.D
 Acq On : 02 Jun 2021 10:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 03 02:17:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	105	-0.03
50 S	Toluene-d8	1.221	1.250	-2.4	109	0.00
51 T	4-Methyl-2-Pentanone	0.555	0.648	-16.8	109	0.00
52 CM	Toluene	0.901	0.969	-7.5#	103	0.00
53 T	t-1,3-Dichloropropene	0.569	0.614	-7.9	102	0.00
54 T	cis-1,3-Dichloropropene	0.611	0.676	-10.6	105	0.00
55 T	1,1,2-Trichloroethane	0.366	0.393	-7.4	102	0.00
56 T	Ethyl methacrylate	0.591	0.661	-11.8	103	0.00
57 T	1,3-Dichloropropane	0.624	0.692	-10.9	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.320	0.319	0.3	98	0.00
59 T	2-Hexanone	0.418	0.498	-19.1	111	0.00
60 T	Dibromochloromethane	0.356	0.369	-3.7	95	0.00
61 T	1,2-Dibromoethane	0.364	0.405	-11.3	103	0.00
62 S	4-Bromofluorobenzene	0.451	0.482	-6.9	113	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.328	0.348	-6.1	102	0.00
65 PM	Chlorobenzene	1.041	1.081	-3.8	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.362	0.376	-3.9	97	0.00
67 C	Ethyl Benzene	1.915	2.035	-6.3#	103	0.00
68 T	m/p-Xylenes	0.724	0.764	-5.5	104	0.00
69 T	o-Xylene	0.704	0.735	-4.4	101	0.00
70 T	Styrene	1.153	1.260	-9.3	103	0.00
71 P	Bromoform	0.250	0.249	0.4	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	4.232	4.305	-1.7	103	0.00
74 T	N-amyl acetate	2.009	2.292	-14.1	113	0.00
75 P	1,1,2,2-Tetrachloroethane	1.403	1.509	-7.6	108	0.00
76 T	1,2,3-Trichloropropane	1.234	1.377	-11.6	108	0.00
77 T	Bromobenzene	0.897	0.907	-1.1	100	0.00
78 T	n-propylbenzene	4.957	5.236	-5.6	105	0.00
79 T	2-Chlorotoluene	3.006	3.039	-1.1	103	0.00
80 T	1,3,5-Trimethylbenzene	3.507	3.619	-3.2	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.521	0.515	1.2	96	0.00
82 T	4-Chlorotoluene	3.464	3.540	-2.2	105	0.00
83 T	tert-Butylbenzene	3.364	3.370	-0.2	99	0.00
84 T	1,2,4-Trimethylbenzene	3.530	3.692	-4.6	103	0.00
85 T	sec-Butylbenzene	4.275	4.476	-4.7	103	0.00
86 T	p-Isopropyltoluene	3.610	3.693	-2.3	101	0.00
87 T	1,3-Dichlorobenzene	1.730	1.789	-3.4	103	0.00
88 T	1,4-Dichlorobenzene	1.788	1.777	0.6	102	0.00
89 T	n-Butylbenzene	3.340	3.609	-8.1	105	0.00
90 T	Hexachloroethane	0.597	0.575	3.7	91	0.00
91 T	1,2-Dichlorobenzene	1.662	1.678	-1.0	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.300	0.317	-5.7	104	0.00
93 T	1,2,4-Trichlorobenzene	1.003	1.040	-3.7	104	0.00
94 T	Hexachlorobutadiene	0.391	0.357	8.7	91	0.00
95 T	Naphthalene	3.910	4.237	-8.4	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060221\
Data File : VX022328.D
Acq On : 02 Jun 2021 10:35
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 03 02:17:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.011	1.027	-1.6	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6