

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029136.D
 Acq On : 02 Jun 2022 15:31
 Operator : JC/MD
 Sample : VX0602WBL01
 Misc : 5.00mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0602WBL01

Quant Time: Jun 03 04:31:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	261873	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	474635	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	451289	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	205612	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	171921	58.219	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	116.440%
35) Dibromofluoromethane	5.385	113	158022	50.124	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.240%
50) Toluene-d8	8.647	98	560553	48.615	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.220%
62) 4-Bromofluorobenzene	11.079	95	218806	49.091	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029136.D
Acq On : 02 Jun 2022 15:31
Operator : JC/MD
Sample : VX0602WBL01
Misc : 5.00mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0602WBL01

Quant Time: Jun 03 04:31:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

