

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029141.D
 Acq On : 02 Jun 2022 17:29
 Operator : JC/MD
 Sample : N3022-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33S-20220525

Quant Time: Jun 03 04:33:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	294805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	540290	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	521205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	240116	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	197272	59.342	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	118.680%
35) Dibromofluoromethane	5.391	113	180862	50.397	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.800%
50) Toluene-d8	8.652	98	639155	48.696	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.400%
62) 4-Bromofluorobenzene	11.085	95	253670	49.997	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029141.D
Acq On : 02 Jun 2022 17:29
Operator : JC/MD
Sample : N3022-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33S-20220525

Quant Time: Jun 03 04:33:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

