

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029142.D
 Acq On : 02 Jun 2022 17:52
 Operator : JC/MD
 Sample : N3022-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB06-20220525

Quant Time: Jun 03 04:33:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	294237	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	546541	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	526095	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	236478	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	198294	59.764	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	119.520%
35) Dibromofluoromethane	5.391	113	180216	49.643	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.280%
50) Toluene-d8	8.653	98	645858	48.644	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.280%
62) 4-Bromofluorobenzene	11.085	95	253048	49.304	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.600%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	22159	16.960	ug/l	98
20) Methylene Chloride	2.788	84	4153	1.363	ug/l	95
25) 2-Butanone	4.556	43	4011	1.903	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029142.D
Acq On : 02 Jun 2022 17:52
Operator : JC/MD
Sample : N3022-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB06-20220525

Quant Time: Jun 03 04:33:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

