

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029143.D
 Acq On : 02 Jun 2022 18:15
 Operator : JC/MD
 Sample : N3022-13DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 08BR-20220525DL

Quant Time: Jun 03 04:33:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	292201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	538306	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	518383	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	239518	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	192925	58.551	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	117.100%
35) Dibromofluoromethane	5.391	113	176974	49.496	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.000%
50) Toluene-d8	8.653	98	627924	48.016	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.040%
62) 4-Bromofluorobenzene	11.085	95	252584	49.967	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.940%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	998	0.370	ug/l	96
20) Methylene Chloride	2.794	84	1513	0.500	ug/l #	82
27) cis-1,2-Dichloroethene	4.495	96	7499	2.116	ug/l	88
44) Trichloroethene	7.129	130	120777	29.612	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029143.D
Acq On : 02 Jun 2022 18:15
Operator : JC/MD
Sample : N3022-13DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
08BR-20220525DL

Quant Time: Jun 03 04:33:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

