

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029144.D
 Acq On : 02 Jun 2022 18:39
 Operator : JC/MD
 Sample : N3022-09DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 91BR-20220525DL

Quant Time: Jun 03 04:33:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	291515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	541859	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	528453	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	248794	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	194126	59.054	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	118.100%
35) Dibromofluoromethane	5.391	113	178894	49.705	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.400%
50) Toluene-d8	8.653	98	644875	48.989	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.980%
62) 4-Bromofluorobenzene	11.085	95	260319	51.159	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.320%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	628	0.248	ug/l	86
4) Vinyl Chloride	1.374	62	45605	15.637	ug/l	100
20) Methylene Chloride	2.794	84	1312	0.435	ug/l #	92
21) trans-1,2-Dichloroethene	3.099	96	2363	0.797	ug/l #	75
27) cis-1,2-Dichloroethene	4.489	96	170370	48.196	ug/l	86
44) Trichloroethene	7.129	130	8566	1.424	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029144.D
Acq On : 02 Jun 2022 18:39
Operator : JC/MD
Sample : N3022-09DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
91BR-20220525DL

Quant Time: Jun 03 04:33:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

