

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029145.D
 Acq On : 02 Jun 2022 19:02
 Operator : JC/MD
 Sample : N3022-12DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 48BR-20220525DL

Quant Time: Jun 03 04:33:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	305362	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	564309	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	547792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	253823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	200453	58.214	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	116.420%
35) Dibromofluoromethane	5.391	113	186778	49.830	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.660%
50) Toluene-d8	8.653	98	665162	48.520	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.040%
62) 4-Bromofluorobenzene	11.079	95	264828	49.975	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.940%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.312	96	901	0.327	ug/l	91
20) Methylene Chloride	2.788	84	1114	0.352	ug/l	90
27) cis-1,2-Dichloroethene	4.489	96	5430	1.466	ug/l	91
44) Trichloroethene	7.129	130	159856	37.575	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029145.D
Acq On : 02 Jun 2022 19:02
Operator : JC/MD
Sample : N3022-12DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
48BR-20220525DL

Quant Time: Jun 03 04:33:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

