

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029146.D
 Acq On : 02 Jun 2022 19:26
 Operator : JC/MD
 Sample : N3022-04 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 45BR-20220525

Quant Time: Jun 03 04:33:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	293338	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	542056	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	524738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	240742	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	193903	58.620	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	117.240%
35) Dibromofluoromethane	5.391	113	180995	50.270	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.540%
50) Toluene-d8	8.653	98	646424	49.089	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.180%
62) 4-Bromofluorobenzene	11.079	95	252797	49.663	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.320%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	2641	0.900	ug/l #	81
20) Methylene Chloride	2.794	84	1097	0.361	ug/l #	86
21) trans-1,2-Dichloroethene	3.093	96	3409	1.143	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	142704	40.118	ug/l	87
44) Trichloroethene	7.129	130	506957	125.693	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029146.D
Acq On : 02 Jun 2022 19:26
Operator : JC/MD
Sample : N3022-04 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
45BR-20220525

Quant Time: Jun 03 04:33:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

