

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029148.D
 Acq On : 02 Jun 2022 20:13
 Operator : JC/MD
 Sample : N3022-15 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP03-20220525

Quant Time: Jun 03 04:34:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	294734	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	539894	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	525294	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	239357	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	191675	57.672	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 115.340%	
35) Dibromofluoromethane	5.391	113	180376	50.299	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.600%	
50) Toluene-d8	8.653	98	642761	49.006	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.020%	
62) 4-Bromofluorobenzene	11.079	95	250823	49.472	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.940%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	1080	0.421	ug/l	95
4) Vinyl Chloride	1.373	62	14996	5.086	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.331	101	1216	0.447	ug/l	96
12) 1,1-Dichloroethene	2.324	96	1869	0.704	ug/l #	89
20) Methylene Chloride	2.788	84	1296	0.425	ug/l #	87
21) trans-1,2-Dichloroethene	3.099	96	3139	1.047	ug/l #	86
27) cis-1,2-Dichloroethene	4.489	96	333076	93.194	ug/l	86
44) Trichloroethene	7.128	130	389469	96.787	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029148.D
Acq On : 02 Jun 2022 20:13
Operator : JC/MD
Sample : N3022-15 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP03-20220525

Quant Time: Jun 03 04:34:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

