

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029156.D
 Acq On : 02 Jun 2022 23:42
 Operator : JC/MD
 Sample : N3164-04 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 23BR-20220531

Quant Time: Jun 03 04:35:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	301785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	553632	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	544127	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	255417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	196837	57.841	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	115.680%
35) Dibromofluoromethane	5.391	113	184242	50.102	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.200%
50) Toluene-d8	8.653	98	660823	49.133	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.260%
62) 4-Bromofluorobenzene	11.079	95	261967	50.388	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.780%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	2590	1.933	ug/l	91
20) Methylene Chloride	2.794	84	885	0.283	ug/l #	91
27) cis-1,2-Dichloroethene	4.501	96	901	0.246	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029156.D
Acq On : 02 Jun 2022 23:42
Operator : JC/MD
Sample : N3164-04 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
23BR-20220531

Quant Time: Jun 03 04:35:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

