

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029158.D
 Acq On : 03 Jun 2022 00:29
 Operator : JC/MD
 Sample : N3164-06 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 28BR-20220531

Quant Time: Jun 03 04:35:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	297822	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	555932	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	538451	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	243803	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	197113	58.693	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	117.380%
35) Dibromofluoromethane	5.391	113	186316	50.456	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.920%
50) Toluene-d8	8.653	98	659885	48.861	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.720%
62) 4-Bromofluorobenzene	11.079	95	255604	48.961	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.920%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	2227	1.684	ug/l	89
20) Methylene Chloride	2.788	84	806	0.261	ug/l	96
37) Ethyl Acetate	4.739	43	3974	0.863	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029158.D
Acq On : 03 Jun 2022 00:29
Operator : JC/MD
Sample : N3164-06 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
28BR-20220531

Quant Time: Jun 03 04:35:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

