

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029164.D
 Acq On : 03 Jun 2022 03:35
 Operator : JC/MD
 Sample : VX0602WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0602WBL02

Quant Time: Jun 03 04:38:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	310240	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	576660	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	563884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	267792	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	198317	56.688	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.380%
35) Dibromofluoromethane	5.391	113	187515	48.956	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.920%
50) Toluene-d8	8.653	98	676722	48.306	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.620%
62) 4-Bromofluorobenzene	11.085	95	274879	50.760	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029164.D
Acq On : 03 Jun 2022 03:35
Operator : JC/MD
Sample : VX0602WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0602WBL02

Quant Time: Jun 03 04:38:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

