

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029173.D
 Acq On : 03 Jun 2022 07:05
 Operator : JC/MD
 Sample : N3164-13 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 37BR-20220531

Quant Time: Jun 03 09:43:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	320715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	592534	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	567995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	259582	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	209385	57.897	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 115.800%			
35) Dibromofluoromethane	5.391	113	194761	49.485	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 98.980%			
50) Toluene-d8	8.653	98	693481	48.176	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 96.360%			
62) 4-Bromofluorobenzene	11.085	95	274592	49.349	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 98.700%			
Target Compounds						
20) Methylene Chloride	2.794	84	855	0.258	ug/l	89
37) Ethyl Acetate	4.733	43	2926	0.596	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029173.D
Acq On : 03 Jun 2022 07:05
Operator : JC/MD
Sample : N3164-13 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
37BR-20220531

Quant Time: Jun 03 09:43:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

