

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029174.D
 Acq On : 03 Jun 2022 07:28
 Operator : JC/MD
 Sample : N3164-14 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 37S-20220531

Quant Time: Jun 03 09:43:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	312527	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	578857	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	561174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	265445	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	204167	57.933	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 115.860%			
35) Dibromofluoromethane	5.391	113	191458	49.795	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.600%			
50) Toluene-d8	8.653	98	687524	48.891	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 97.780%			
62) 4-Bromofluorobenzene	11.079	95	271815	50.004	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 100.000%			
Target Compounds						
16) Acetone	2.380	43	2233	1.609	ug/l	94
20) Methylene Chloride	2.794	84	824	0.255	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
Data File : VX029174.D
Acq On : 03 Jun 2022 07:28
Operator : JC/MD
Sample : N3164-14 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
37S-20220531

Quant Time: Jun 03 09:43:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

