

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060222\
 Data File : VX029175.D
 Acq On : 03 Jun 2022 07:51
 Operator : JC/MD
 Sample : N3164-15 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 38BR-20220531

Quant Time: Jun 03 09:43:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	315517	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	587739	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	569000	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	257261	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	206888	58.149	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	116.300%
35) Dibromofluoromethane	5.391	113	192098	49.207	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.420%
50) Toluene-d8	8.653	98	689168	48.267	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.540%
62) 4-Bromofluorobenzene	11.079	95	272627	49.396	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.800%
Target Compounds						
4) Vinyl Chloride	1.374	62	18707	5.926	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 03 09:43:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

