

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060222\  
 Data File : VX029161.D  
 Acq On : 03 Jun 2022 01:38  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 29 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC

Manual Integrations  
 APPROVED

Reviewed By : Mahesh Dadoda 06/03/2022  
 Supervised By : Semsettin Yesilyurt 06/03/2022

Quant Time: Jun 03 04:36:59 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X053122W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 01 04:45:42 2022  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                        |                |      |                     |         |        |          |
| Internal Standards           |                |      |                     |         |        |          |
| 1) Pentafluorobenzene        | 5.556          | 168  | 241640              | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763          | 114  | 417979              | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 377119              | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 188349              | 50.000  | ug/l   | 0.00     |
|                              |                |      |                     |         |        |          |
| System Monitoring Compounds  |                |      |                     |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.958          | 65   | 157822              | 57.920  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 61 - 141 |      | Recovery = 115.840% |         |        |          |
| 35) Dibromofluoromethane     | 5.391          | 113  | 143691              | 51.756  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 69 - 133 |      | Recovery = 103.520% |         |        |          |
| 50) Toluene-d8               | 8.653          | 98   | 524211              | 51.625  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 65 - 126 |      | Recovery = 103.260% |         |        |          |
| 62) 4-Bromofluorobenzene     | 11.085         | 95   | 201090              | 51.232  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 58 - 135 |      | Recovery = 102.460% |         |        |          |
|                              |                |      |                     |         |        |          |
| Target Compounds             |                |      |                     |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.166          | 85   | 129988              | 61.869  | ug/l   | 97       |
| 3) Chloromethane             | 1.288          | 50   | 117605              | 53.495  | ug/l   | 99       |
| 4) Vinyl Chloride            | 1.374          | 62   | 133284              | 55.131  | ug/l   | 98       |
| 5) Bromomethane              | 1.605          | 94   | 60436               | 54.696  | ug/l   | 100      |
| 6) Chloroethane              | 1.678          | 64   | 85889               | 49.176  | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 1.886          | 101  | 213607              | 56.850  | ug/l   | 98       |
| 8) Diethyl Ether             | 2.136          | 74   | 78766               | 54.698  | ug/l   | 97       |
| 9) 1,1,2-Trichlorotrifluo... | 2.331          | 101  | 130085              | 58.375  | ug/l   | 99       |
| 10) Methyl Iodide            | 2.453          | 142  | 165826              | 59.797  | ug/l   | 98       |
| 11) Tert butyl alcohol       | 2.989          | 59   | 187076              | 296.371 | ug/l   | 97       |
| 12) 1,1-Dichloroethene       | 2.319          | 96   | 126969              | 58.299  | ug/l   | 95       |
| 13) Acrolein                 | 2.239          | 56   | 63286               | 221.278 | ug/l   | 98       |
| 14) Allyl chloride           | 2.666          | 41   | 189855              | 58.045  | ug/l   | 97       |
| 15) Acrylonitrile            | 3.068          | 53   | 401263              | 308.737 | ug/l   | 99       |
| 16) Acetone                  | 2.379          | 43   | 320094              | 298.325 | ug/l   | 98       |
| 17) Carbon Disulfide         | 2.514          | 76   | 301215              | 50.426  | ug/l   | 100      |
| 18) Methyl Acetate           | 2.703          | 43   | 176627              | 63.649  | ug/l   | 98       |
| 19) Methyl tert-butyl Ether  | 3.117          | 73   | 452089              | 61.182  | ug/l   | 99       |
| 20) Methylene Chloride       | 2.788          | 84   | 149502              | 59.762  | ug/l   | 99       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 141036              | 57.401  | ug/l   | 96       |
| 22) Diisopropyl ether        | 3.763          | 45   | 408026              | 62.369  | ug/l   | 95       |
| 23) Vinyl Acetate            | 3.721          | 43   | 1729061             | 316.463 | ug/l   | 100      |
| 24) 1,1-Dichloroethane       | 3.611          | 63   | 246601              | 60.925  | ug/l   | 99       |
| 25) 2-Butanone               | 4.556          | 43   | 529013              | 305.641 | ug/l   | 99       |
| 26) 2,2-Dichloropropane      | 4.477          | 77   | 143275              | 43.146  | ug/l   | 97       |
| 27) cis-1,2-Dichloroethene   | 4.489          | 96   | 166114              | 56.691  | ug/l   | 92       |
| 28) Bromochloromethane       | 4.903          | 49   | 94512               | 58.031  | ug/l   | 97       |
| 29) Tetrahydrofuran          | 5.007          | 42   | 344489              | 310.368 | ug/l   | 97       |
| 30) Chloroform               | 5.098          | 83   | 268202              | 59.770  | ug/l   | 99       |
| 31) Cyclohexane              | 5.476          | 56   | 214778              | 60.630  | ug/l   | 97       |
| 32) 1,1,1-Trichloroethane    | 5.385          | 97   | 234560              | 57.727  | ug/l   | 98       |
| 36) 1,1-Dichloropropene      | 5.696          | 75   | 193324              | 54.117  | ug/l   | 99       |
| 37) Ethyl Acetate            | 4.721          | 43   | 198929              | 57.459  | ug/l   | 99       |
| 38) Carbon Tetrachloride     | 5.678          | 117  | 204507              | 50.093  | ug/l   | 99       |
| 39) Methylcyclohexane        | 7.385          | 83   | 235303              | 53.285  | ug/l   | 99       |
| 40) Benzene                  | 6.037          | 78   | 585190              | 55.308  | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060222\  
 Data File : VX029161.D  
 Acq On : 03 Jun 2022 01:38  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 29 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022  
 Supervised By :Semsettin Yesilyurt 06/03/2022

Quant Time: Jun 03 04:36:59 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X053122W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 01 04:45:42 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 41) Methacrylonitrile         | 4.922  | 41   | 107814   | 57.656   | ug/l   | 96       |
| 42) 1,2-Dichloroethane        | 6.086  | 62   | 197121   | 54.939   | ug/l   | 98       |
| 43) Isopropyl Acetate         | 6.342  | 43   | 299798   | 55.351   | ug/l   | 99       |
| 44) Trichloroethene           | 7.129  | 130  | 163428   | 52.133   | ug/l   | 93       |
| 45) 1,2-Dichloropropane       | 7.433  | 63   | 144740   | 56.766   | ug/l   | 99       |
| 46) Dibromomethane            | 7.586  | 93   | 104144   | 52.431   | ug/l   | 95       |
| 47) Bromodichloromethane      | 7.824  | 83   | 202961   | 51.832   | ug/l   | 99       |
| 48) Methyl methacrylate       | 7.696  | 41   | 150620   | 57.169   | ug/l   | 95       |
| 49) 1,4-Dioxane               | 7.671  | 88   | 84569    | 1048.993 | ug/l   | 97       |
| 51) 4-Methyl-2-Pentanone      | 8.573  | 43   | 1006205  | 279.312  | ug/l   | 99       |
| 52) Toluene                   | 8.720  | 92   | 387328   | 54.803   | ug/l   | 97       |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 201766   | 50.136   | ug/l   | 98       |
| 54) cis-1,3-Dichloropropene   | 8.366  | 75   | 225125   | 51.127   | ug/l   | 97       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 159084   | 54.094   | ug/l   | 99       |
| 56) Ethyl methacrylate        | 9.116  | 69   | 235462   | 56.406   | ug/l   | 97       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 256532   | 54.726   | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 8.244  | 63   | 549513   | 277.791  | ug/l   | 96       |
| 59) 2-Hexanone                | 9.433  | 43   | 784203   | 279.717  | ug/l   | 98       |
| 60) Dibromochloromethane      | 9.525  | 129  | 160747   | 48.026   | ug/l   | 96       |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 169798   | 53.383   | ug/l   | 99       |
| 64) Tetrachloroethene         | 9.275  | 164  | 144586   | 50.919   | ug/l   | 96       |
| 65) Chlorobenzene             | 10.079 | 112  | 408287   | 52.701   | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 152240   | 50.978   | ug/l   | 99       |
| 67) Ethyl Benzene             | 10.195 | 91   | 732855   | 56.299   | ug/l   | 99       |
| 68) m/p-Xylenes               | 10.305 | 106  | 574857   | 109.963  | ug/l   | 99       |
| 69) o-Xylene                  | 10.646 | 106  | 283764   | 54.923   | ug/l   | 100      |
| 70) Styrene                   | 10.658 | 104  | 479152   | 55.725   | ug/l   | 99       |
| 71) Bromoform                 | 10.805 | 173  | 112638   | 43.688   | ug/l # | 99       |
| 73) Isopropylbenzene          | 10.963 | 105  | 737032   | 58.230   | ug/l   | 99       |
| 74) N-amyl acetate            | 10.847 | 43   | 252009   | 61.489   | ug/l   | 99       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 245522   | 57.541   | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 11.244 | 75   | 217232m  | 56.895   | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 174171   | 51.593   | ug/l   | 97       |
| 78) n-propylbenzene           | 11.305 | 91   | 837640   | 59.502   | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 512606   | 57.608   | ug/l   | 98       |
| 80) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 629408   | 58.085   | ug/l   | 100      |
| 81) trans-1,4-Dichloro-2-b... | 11.018 | 75   | 59423    | 45.660   | ug/l   | 90       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 586936   | 57.622   | ug/l   | 99       |
| 83) tert-Butylbenzene         | 11.719 | 119  | 613439   | 55.389   | ug/l   | 95       |
| 84) 1,2,4-Trimethylbenzene    | 11.756 | 105  | 630951   | 58.334   | ug/l   | 99       |
| 85) sec-Butylbenzene          | 11.890 | 105  | 766851   | 58.978   | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 643681   | 56.623   | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 332983   | 52.335   | ug/l   | 97       |
| 88) 1,4-Dichlorobenzene       | 12.042 | 146  | 336018   | 51.251   | ug/l   | 97       |
| 89) n-Butylbenzene            | 12.335 | 91   | 532033   | 57.674   | ug/l   | 100      |
| 90) Hexachloroethane          | 12.542 | 117  | 100996   | 48.864   | ug/l   | 91       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 339480   | 54.166   | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 55417    | 58.168   | ug/l   | 97       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 205809   | 52.658   | ug/l   | 98       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 83184    | 51.001   | ug/l   | 93       |
| 95) Naphthalene               | 13.780 | 128  | 763542   | 61.092   | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 209319   | 54.234   | ug/l   | 95       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060222\  
Data File : VX029161.D  
Acq On : 03 Jun 2022 01:38  
Operator : JC/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 29 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022

Quant Time: Jun 03 04:36:59 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X053122W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jun 01 04:45:42 2022  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|


-----  
(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.556 min Scan# 71  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:168 Resp: 241640  
Ion Ratio Lower Upper  
168 100  
99 59.2 41.7 62.5

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#2  
Dichlorodifluoromethane  
Concen: 61.869 ug/l  
RT: 1.166 min Scan# 13  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Tgt Ion: 85 Resp: 129988  
Ion Ratio Lower Upper  
85 100  
87 30.8 16.2 48.6





#3  
Chloromethane  
Concen: 53.495 ug/l  
RT: 1.288 min Scan# 33  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38



Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 50 Resp: 117609  
Ion Ratio Lower Upper  
50 100  
52 33.8 26.4 39.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#4  
Vinyl Chloride  
Concen: 55.131 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38

Tgt Ion: 62 Resp: 133284  
Ion Ratio Lower Upper  
62 100  
64 33.0 25.5 38.3









#5  
Bromomethane  
Concen: 54.696 ug/l  
RT: 1.605 min Scan# 84  
Delta R.T. 0.006 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion: 94 Resp: 60436  
Ion Ratio Lower Upper  
94 100  
96 94.3 75.3 112.9




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#6  
Chloroethane  
Concen: 49.176 ug/l  
RT: 1.678 min Scan# 97  
Delta R.T. -0.001 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38


Tgt Ion: 64 Resp: 85889  
Ion Ratio Lower Upper  
64 100  
66 31.9 25.3 37.9





#7  
Trichlorofluoromethane  
Concen: 56.850 ug/l  
RT: 1.886 min Scan# 11  
Delta R.T. 0.006 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:101 Resp: 21360  
Ion Ratio Lower Upper  
101 100  
103 64.9 50.5 75.7




Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#8  
Diethyl Ether  
Concen: 54.698 ug/l  
RT: 2.136 min Scan# 172  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38

Tgt Ion: 74 Resp: 78766  
Ion Ratio Lower Upper  
74 100  
45 80.8 42.0 126.0








#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 58.375 ug/l  
RT: 2.331 min Scan# 203  
Delta R.T. 0.006 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:101 Resp: 130085  
Ion Ratio Lower Upper  
101 100  
85 42.3 33.5 50.3  
151 76.7 61.8 92.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#10  
Methyl Iodide  
Concen: 59.797 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38


Tgt Ion:142 Resp: 165826  
Ion Ratio Lower Upper  
142 100  
127 43.7 33.9 50.9  
141 14.0 11.4 17.2





#11  
Tert butyl alcohol  
Concen: 296.371 ug/l  
RT: 2.989 min Scan# 312  
Delta R.T. 0.018 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 59 Resp: 187070  
Ion Ratio Lower Upper  
59 100  
57 9.1 8.3 12.5




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#12  
1,1-Dichloroethene  
Concen: 58.299 ug/l  
RT: 2.319 min Scan# 202  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38


Tgt Ion: 96 Resp: 126969  
Ion Ratio Lower Upper  
96 100  
61 147.2 124.6 187.0  
98 65.1 51.0 76.4





#13  
Acrolein  
Concen: 221.278 ug/l  
RT: 2.239 min Scan# 189  
Delta R.T. 0.006 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38



Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 56 Resp: 63280  
Ion Ratio Lower Upper  
56 100  
55 66.8 55.0 82.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#14  
Allyl chloride  
Concen: 58.045 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. 0.006 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38


Tgt Ion: 41 Resp: 189855  
Ion Ratio Lower Upper  
41 100  
39 61.4 50.4 75.6  
76 38.3 32.7 49.1





#15  
Acrylonitrile  
Concen: 308.737 ug/l  
RT: 3.068 min Scan# 31  
Delta R.T. 0.006 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 53 Resp: 40126  
Ion Ratio Lower Upper  
53 100  
52 81.7 65.7 98.5  
51 34.9 28.4 42.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#16  
Acetone  
Concen: 298.325 ug/l  
RT: 2.379 min Scan# 212  
Delta R.T. -0.001 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38


Tgt Ion: 43 Resp: 320094  
Ion Ratio Lower Upper  
43 100  
58 34.8 28.8 43.2





#17  
Carbon Disulfide  
Concen: 50.426 ug/l  
RT: 2.514 min Scan# 21  
Delta R.T. 0.006 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 76 Resp: 301215  
Ion Ratio Lower Upper  
76 100  
78 9.1 7.2 10.8




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#18  
Methyl Acetate  
Concen: 63.649 ug/l  
RT: 2.703 min Scan# 265  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38

Tgt Ion: 43 Resp: 176627  
Ion Ratio Lower Upper  
43 100  
74 27.6 22.8 34.2









#19  
Methyl tert-butyl Ether  
Concen: 61.182 ug/l  
RT: 3.117 min Scan# 311  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion: 73 Resp: 452089  
Ion Ratio Lower Upper  
73 100  
57 20.7 16.8 25.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#20  
Methylene Chloride  
Concen: 59.762 ug/l  
RT: 2.788 min Scan# 279  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38


Tgt Ion: 84 Resp: 149502  
Ion Ratio Lower Upper  
84 100  
49 107.6 85.5 128.3  
51 32.6 26.6 39.8  
86 66.6 51.5 77.3





#21  
trans-1,2-Dichloroethene  
Concen: 57.401 ug/l  
RT: 3.093 min Scan# 31  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 96 Resp: 141036  
Ion Ratio Lower Upper  
96 100  
61 125.9 105.4 158.2  
98 64.4 50.0 75.0




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#22  
Diisopropyl ether  
Concen: 62.369 ug/l  
RT: 3.763 min Scan# 439  
Delta R.T. -0.001 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38


Tgt Ion: 45 Resp: 408026  
Ion Ratio Lower Upper  
45 100  
43 69.1 50.8 76.2  
87 34.0 26.6 39.8  
59 12.9 10.5 15.7





#23  
 Vinyl Acetate  
 Concen: 316.463 ug/l  
 RT: 3.721 min Scan# 41  
 Delta R.T. -0.000 min  
 Lab File: VX029161.D  
 Acq: 03 Jun 2022 01:38




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC




Tgt Ion: 43 Resp: 172906  
 Ion Ratio Lower Upper  
 43 100  
 86 13.5 10.8 16.2

Manual Integrations  
 APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
 Supervised By :Semsettin Yesilyurt 06/03/2022




#24  
 1,1-Dichloroethane  
 Concen: 60.925 ug/l  
 RT: 3.611 min Scan# 414  
 Delta R.T. -0.000 min  
 Lab File: VX029161.D  
 Acq: 03 Jun 2022 01:38



Tgt Ion: 63 Resp: 246601  
 Ion Ratio Lower Upper  
 63 100  
 98 8.2 3.9 11.7  
 100 5.0 2.4 7.2








#25  
2-Butanone  
Concen: 305.641 ug/l  
RT: 4.556 min Scan# 569  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 43 Resp: 529011  
Ion Ratio Lower Upper  
43 100  
72 30.0 24.4 36.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#26  
2,2-Dichloropropane  
Concen: 43.146 ug/l  
RT: 4.477 min Scan# 556  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38

Tgt Ion: 77 Resp: 143275  
Ion Ratio Lower Upper  
77 100  
97 25.3 11.8 35.4





#27

cis-1,2-Dichloroethene

Concen: 56.691 ug/l

RT: 4.489 min Scan# 558

Delta R.T. -0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 96 Resp: 166114

Ion Ratio Lower Upper

96 100





61 128.1 0.0 283.8

98 65.3 0.0 129.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#28

Bromochloromethane

Concen: 58.031 ug/l

RT: 4.903 min Scan# 626

Delta R.T. 0.006 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38




Tgt Ion: 49 Resp: 94512


Ion Ratio Lower Upper

49 100

129 2.3 0.0 4.4

130 91.7 70.7 106.1





#29

Tetrahydrofuran

Concen: 310.368 ug/l

RT: 5.007 min Scan# 643

Delta R.T. 0.000 min

Lab File: VX029161.D


Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC






|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 42    | Resp: | 344489 |
| Ion      | Ratio | Lower | Upper  |
| 42       | 100   |       |        |
| 72       | 53.3  | 44.1  | 66.1   |
| 71       | 48.5  | 40.5  | 60.7   |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#30

Chloroform

Concen: 59.770 ug/l




RT: 5.098 min Scan# 658


Delta R.T. -0.001 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 83    | Resp: | 268202 |
| Ion      | Ratio | Lower | Upper  |
| 83       | 100   |       |        |
| 85       | 66.9  | 52.7  | 79.1   |





#31  
Cyclohexane  
Concen: 60.630 ug/l  
RT: 5.476 min Scan# 719  
Delta R.T. 0.005 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 56 Resp: 214778  
Ion Ratio Lower Upper  
56 100  
69 33.0 26.7 40.1  
84 99.8 77.2 115.8




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#32  
1,1,1-Trichloroethane  
Concen: 57.727 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38

Tgt Ion: 97 Resp: 234560  
Ion Ratio Lower Upper  
97 100  
99 64.2 51.8 77.6  
61 39.6 34.0 51.0





#33

1,2-Dichloroethane-d4

Concen: 57.920 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 15782

Ion Ratio Lower Upper

65 100





67 53.5 0.0 107.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022

Supervised By :Semsettin Yesilyurt 06/03/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38




Tgt Ion:114 Resp: 417979


Ion Ratio Lower Upper

114 100

63 19.0 0.0 36.8


88 16.1 0.0 32.4





#35  
Dibromofluoromethane  
Concen: 51.756 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 113 Resp: 14369  
Ion Ratio Lower Upper  
113 100  
111 102.0 82.8 124.2  
192 19.2 16.9 25.3




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#36  
1,1-Dichloropropene  
Concen: 54.117 ug/l  
RT: 5.696 min Scan# 756  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38

Tgt Ion: 75 Resp: 193324  
Ion Ratio Lower Upper  
75 100  
110 37.5 18.3 54.8  
77 31.0 24.8 37.2





#37

Ethyl Acetate

Concen: 57.459 ug/l

RT: 4.721 min Scan# 596

Delta R.T. 0.006 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 198929

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 43  | 100   |       |       |
| 61  | 15.7  | 12.6  | 19.0  |
| 70  | 13.6  | 10.2  | 15.2  |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#38

Carbon Tetrachloride

Concen: 50.093 ug/l

RT: 5.678 min Scan# 753




Delta R.T. -0.000 min


Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Tgt Ion:117 Resp: 204507

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 117 | 100   |       |       |
| 119 | 95.4  | 77.3  | 115.9 |
| 121 | 31.0  | 24.2  | 36.2  |





#39

Methylcyclohexane

Concen: 53.285 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 23530

Ion Ratio Lower Upper

83 100





55 68.4 54.6 81.8

98 48.2 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#40

Benzene

Concen: 55.308 ug/l

RT: 6.037 min Scan# 812

Delta R.T. -0.007 min

Lab File: VX029161.D




Acq: 03 Jun 2022 01:38

Tgt Ion: 78 Resp: 585190


Ion Ratio Lower Upper

78 100

77 22.9 19.3 28.9








#41  
Methacrylonitrile  
Concen: 57.656 ug/l  
RT: 4.922 min Scan# 61  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 41 Resp: 107814  
Ion Ratio Lower Upper  
41 100  
39 53.1 43.0 64.4  
67 81.8 70.2 105.4  
52 30.9 26.0 39.0

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#42  
1,2-Dichloroethane  
Concen: 54.939 ug/l  
RT: 6.086 min Scan# 820  
Delta R.T. -0.006 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Tgt Ion: 62 Resp: 197121  
Ion Ratio Lower Upper  
62 100  
98 10.8 0.0 20.4





#43  
Isopropyl Acetate  
Concen: 55.351 ug/l  
RT: 6.342 min Scan# 862  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 43 Resp: 299798  
Ion Ratio Lower Upper  
43 100  
61 22.8 18.5 27.7  
87 16.6 13.0 19.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#44  
Trichloroethene  
Concen: 52.133 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38

Tgt Ion:130 Resp: 163428  
Ion Ratio Lower Upper  
130 100  
95 101.0 0.0 188.8





#45

1,2-Dichloropropane

Concen: 56.766 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.001 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 144740

Ion Ratio Lower Upper





63 100

65 31.6 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#46

Dibromomethane

Concen: 52.431 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38




Tgt Ion: 93 Resp: 104144


Ion Ratio Lower Upper

93 100

95 85.4 65.8 98.6


174 100.9 85.4 128.2





#47  
Bromodichloromethane  
Concen: 51.832 ug/l  
RT: 7.824 min Scan# 1105  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 83 Resp: 20296  
Ion Ratio Lower Upper  
83 100  
85 64.6 52.2 78.2  
127 8.8 6.8 10.2

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#48  
Methyl methacrylate  
Concen: 57.169 ug/l  
RT: 7.696 min Scan# 1084  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Tgt Ion: 41 Resp: 150620  
Ion Ratio Lower Upper  
41 100  
69 96.0 82.5 123.7  
39 53.2 43.8 65.8





#49  
1,4-Dioxane  
Concen: 1048.993 ug/l  
RT: 7.671 min Scan# 1078  
Delta R.T. 0.012 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 88 Resp: 84569  
Ion Ratio Lower Upper  
88 100  
43 26.2 21.9 32.9  
58 64.6 53.7 80.5




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#50  
Toluene-d8  
Concen: 51.625 ug/l  
RT: 8.653 min Scan# 1241  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38


Tgt Ion: 98 Resp: 524211  
Ion Ratio Lower Upper  
98 100  
100 67.4 52.6 78.8





#51  
4-Methyl-2-Pentanone  
Concen: 279.312 ug/l  
RT: 8.573 min Scan# 11  
Delta R.T. -0.001 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 43 Resp: 100620  
Ion Ratio Lower Upper  
43 100  
58 44.8 36.1 54.1

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#52  
Toluene  
Concen: 54.803 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Tgt Ion: 92 Resp: 387328  
Ion Ratio Lower Upper  
92 100  
91 167.2 137.5 206.3





#53  
 t-1,3-Dichloropropene  
 Concen: 50.136 ug/l  
 RT: 8.982 min Scan# 1295  
 Delta R.T. -0.000 min  
 Lab File: VX029161.D  
 Acq: 03 Jun 2022 01:38




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC




Tgt Ion: 75 Resp: 201760  
 Ion Ratio Lower Upper  
 75 100  
 77 30.3 25.4 38.0

Manual Integrations  
 APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
 Supervised By :Semsettin Yesilyurt 06/03/2022




#54  
 cis-1,3-Dichloropropene  
 Concen: 51.127 ug/l  
 RT: 8.366 min Scan# 1194  
 Delta R.T. 0.000 min  
 Lab File: VX029161.D  
 Acq: 03 Jun 2022 01:38




Tgt Ion: 75 Resp: 225125  
 Ion Ratio Lower Upper  
 75 100  
 77 30.5 24.8 37.2  
 39 39.7 29.6 44.4





#55  
1,1,2-Trichloroethane  
Concen: 54.094 ug/l  
RT: 9.153 min Scan# 1323  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 97 Resp: 159084  
Ion Ratio Lower Upper  
97 100  
83 83.0 66.2 99.4  
85 53.3 42.9 64.3  
99 60.6 49.0 73.6

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#56  
Ethyl methacrylate  
Concen: 56.406 ug/l  
RT: 9.116 min Scan# 1317  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38



Tgt Ion: 69 Resp: 235462  
Ion Ratio Lower Upper  
69 100  
41 56.6 43.8 65.6  
39 29.1 22.2 33.2







#57

1,3-Dichloropropane

Concen: 54.726 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. 0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 25653

Ion Ratio Lower Upper





76 100

78 32.5 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#58

2-Chloroethyl Vinyl ether

Concen: 277.791 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX029161.D




Acq: 03 Jun 2022 01:38


Tgt Ion: 63 Resp: 549513

Ion Ratio Lower Upper

63 100


106 32.0 23.9 35.9





#59  
2-Hexanone  
Concen: 279.717 ug/l  
RT: 9.433 min Scan# 1369  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 43 Resp: 78420  
Ion Ratio Lower Upper  
43 100  
58 62.3 31.9 95.7

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#60  
Dibromochloromethane  
Concen: 48.026 ug/l  
RT: 9.525 min Scan# 1384  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Tgt Ion:129 Resp: 160747  
Ion Ratio Lower Upper  
129 100  
127 76.6 40.3 120.8





#61  
1,2-Dibromoethane  
Concen: 53.383 ug/l  
RT: 9.610 min Scan# 1398  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:107 Resp: 169798  
Ion Ratio Lower Upper  
107 100  
109 93.5 74.4 111.6

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#62  
4-Bromofluorobenzene  
Concen: 51.232 ug/l  
RT: 11.085 min Scan# 1640  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Tgt Ion: 95 Resp: 201090  
Ion Ratio Lower Upper  
95 100  
174 74.4 0.0 163.0  
176 71.2 0.0 155.0





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:117 Resp: 377119  
Ion Ratio Lower Upper  
117 100  
82 55.4 41.8 62.8  
119 32.3 24.8 37.2

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#64  
Tetrachloroethene  
Concen: 50.919 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Tgt Ion:164 Resp: 144586  
Ion Ratio Lower Upper  
164 100  
166 123.9 101.2 151.8  
129 100.3 73.0 109.4  
131 93.0 74.2 111.2





#65  
Chlorobenzene  
Concen: 52.701 ug/l  
RT: 10.079 min Scan# 1475  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:112 Resp: 40828  
Ion Ratio Lower Upper  
112 100  
114 31.6 25.8 38.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#66  
1,1,1,2-Tetrachloroethane  
Concen: 50.978 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38



Tgt Ion:131 Resp: 152240  
Ion Ratio Lower Upper  
131 100  
133 95.3 48.3 144.8  
119 65.0 32.6 98.0





#67

Ethyl Benzene

Concen: 56.299 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 91 Resp: 73285

Ion Ratio Lower Upper





91 100

106 31.1 24.6 36.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#68

m/p-Xylenes

Concen: 109.963 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. -0.000 min

Lab File: VX029161.D




Acq: 03 Jun 2022 01:38


Tgt Ion:106 Resp: 574857

Ion Ratio Lower Upper

106 100

91 199.0 160.4 240.6





#69

o-Xylene

Concen: 54.923 ug/l

RT: 10.646 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:106 Resp: 283764

Ion Ratio Lower Upper

106 100





91 212.4 106.1 318.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022

Supervised By :Semsettin Yesilyurt 06/03/2022



#70

Styrene

Concen: 55.725 ug/l

RT: 10.658 min Scan# 1570

Delta R.T. -0.001 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38




Tgt Ion:104 Resp: 479152


Ion Ratio Lower Upper

104 100

78 48.9 38.6 57.8

103 53.7 43.8 65.8





#71

Bromoform

Concen: 43.688 ug/l

RT: 10.805 min Scan# 1594

Delta R.T. -0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:173 Resp: 112638

Ion Ratio Lower Upper

173 100

175 49.0 24.1 72.2





254 0.0 0.1 0.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022

Supervised By :Semsettin Yesilyurt 06/03/2022



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38




Tgt Ion:152 Resp: 188349

Ion Ratio Lower Upper


152 100

115 88.0 37.5 112.7

150 176.3 0.0 338.2








#73  
Isopropylbenzene  
Concen: 58.230 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:105 Resp: 737033  
Ion Ratio Lower Upper  
105 100  
120 26.4 13.5 40.5

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#74  
N-ethyl acetate  
Concen: 61.489 ug/l  
RT: 10.847 min Scan# 1601  
Delta R.T. -0.001 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Tgt Ion: 43 Resp: 252009  
Ion Ratio Lower Upper  
43 100  
70 52.7 42.5 63.7  
55 28.1 23.1 34.7  
61 28.2 22.8 34.2





#75  
1,1,2,2-Tetrachloroethane  
Concen: 57.541 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 83 Resp: 24552  
Ion Ratio Lower Upper  
83 100  
131 10.7 5.1 15.3  
85 65.1 32.6 97.7

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#76  
1,2,3-Trichloropropane  
Concen: 56.895 ug/l m  
RT: 11.244 min Scan# 1666  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38



Tgt Ion: 75 Resp: 217232  
Ion Ratio Lower Upper  
75 100  
77 37.8 19.4 58.2





#77

Bromobenzene

Concen: 51.593 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. 0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:156 Resp: 17417  
 Ion Ratio Lower Upper  
 156 100  
 77 160.4 78.5 235.3  
 158 96.3 50.2 150.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022  
 Supervised By :Semsettin Yesilyurt 06/03/2022



#78

n-propylbenzene

Concen: 59.502 ug/l




RT: 11.305 min Scan# 1676


Delta R.T. -0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Tgt Ion: 91 Resp: 837640  
 Ion Ratio Lower Upper  
 91 100  
 120 23.5 11.6 34.8





#79

2-Chlorotoluene

Concen: 57.608 ug/l

RT: 11.366 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 91 Resp: 512600

Ion Ratio Lower Upper

91 100





126 33.7 17.3 51.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022

Supervised By :Semsettin Yesilyurt 06/03/2022



#80

1,3,5-Trimethylbenzene

Concen: 58.085 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX029161.D




Acq: 03 Jun 2022 01:38


Tgt Ion:105 Resp: 629408

Ion Ratio Lower Upper

105 100


120 48.5 24.3 72.9





#81  
trans-1,4-Dichloro-2-butene  
Concen: 45.660 ug/l  
RT: 11.018 min Scan# 1629  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion: 75 Resp: 5942  
Ion Ratio Lower Upper  
75 100  
53 98.5 69.0 103.4  
89 49.5 37.0 55.6

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#82  
4-Chlorotoluene  
Concen: 57.622 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Tgt Ion: 91 Resp: 586936  
Ion Ratio Lower Upper  
91 100  
126 30.3 14.9 44.5





#83  
tert-Butylbenzene  
Concen: 55.389 ug/l  
RT: 11.719 min Scan# 11744  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:119 Resp: 613439  
Ion Ratio Lower Upper  
119 100  
91 58.5 26.9 80.6  
134 23.3 11.4 34.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#84  
1,2,4-Trimethylbenzene  
Concen: 58.334 ug/l  
RT: 11.756 min Scan# 1750  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38


Tgt Ion:105 Resp: 630951  
Ion Ratio Lower Upper  
105 100  
120 44.6 22.6 67.8





#85  
 sec-Butylbenzene  
 Concen: 58.978 ug/l  
 RT: 11.890 min Scan# 1772  
 Delta R.T. -0.000 min  
 Lab File: VX029161.D  
 Acq: 03 Jun 2022 01:38




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050EC



Tgt Ion:105 Resp: 76685  
 Ion Ratio Lower Upper  
 105 100  
 134 20.2 10.2 30.4




Manual Integrations  
 APPROVED


Reviewed By :Mahesh Dadoda 06/03/2022  
 Supervised By :Semsettin Yesilyurt 06/03/2022



#86  
 p-Isopropyltoluene  
 Concen: 56.623 ug/l  
 RT: 12.012 min Scan# 1792  
 Delta R.T. -0.000 min  
 Lab File: VX029161.D  
 Acq: 03 Jun 2022 01:38

Tgt Ion:119 Resp: 643681  
 Ion Ratio Lower Upper  
 119 100  
 134 26.1 13.0 39.0  
 91 23.9 11.3 33.9





#87

1,3-Dichlorobenzene

Concen: 52.335 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:146 Resp: 33298

Ion Ratio Lower Upper

146 100

111 42.7 19.7 59.1





148 63.7 31.6 94.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022

Supervised By :Semsettin Yesilyurt 06/03/2022



#88

1,4-Dichlorobenzene

Concen: 51.251 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.001 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38




Tgt Ion:146 Resp: 336018

Ion Ratio Lower Upper


146 100

111 41.8 19.2 57.6

148 64.0 32.3 96.9







#89

n-Butylbenzene

Concen: 57.674 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 91 Resp: 53203

Ion Ratio Lower Upper

91 100

92 55.4 27.8 83.3





134 27.1 13.3 39.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022

Supervised By :Semsettin Yesilyurt 06/03/2022



#90

Hexachloroethane

Concen: 48.864 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX029161.D




Acq: 03 Jun 2022 01:38


Tgt Ion:117 Resp: 100996

Ion Ratio Lower Upper

117 100


201 71.4 39.5 118.5





#91  
1,2-Dichlorobenzene  
Concen: 54.166 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050EC




Tgt Ion:146 Resp: 339480  
Ion Ratio Lower Upper  
146 100  
111 43.7 21.0 63.0  
148 63.2 32.6 97.7

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 58.168 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. -0.000 min  
Lab File: VX029161.D  
Acq: 03 Jun 2022 01:38



Tgt Ion: 75 Resp: 55417  
Ion Ratio Lower Upper  
75 100  
155 90.8 47.0 141.2  
157 112.6 57.9 173.6





#93

1,2,4-Trichlorobenzene

Concen: 52.658 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. -0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:180 Resp: 205809

Ion Ratio Lower Upper

180 100

182 95.7 47.7 143.1





145 33.6 15.2 45.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022

Supervised By :Semsettin Yesilyurt 06/03/2022



#94

Hexachlorobutadiene

Concen: 51.001 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38




Tgt Ion:225 Resp: 83184


Ion Ratio Lower Upper

225 100

223 62.4 34.2 102.5

227 62.6 33.9 101.7





#95

Naphthalene

Concen: 61.092 ug/l

RT: 13.780 min Scan# 2082

Delta R.T. -0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 76354

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 128 | 100   |       |       |
| 127 | 12.7  | 10.2  | 15.4  |
| 129 | 11.1  | 8.7   | 13.1  |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/03/2022  
Supervised By :Semsettin Yesilyurt 06/03/2022



#96

1,2,3-Trichlorobenzene

Concen: 54.234 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX029161.D

Acq: 03 Jun 2022 01:38

Tgt Ion:180 Resp: 209319

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 180 | 100   |       |       |
| 182 | 95.3  | 50.0  | 149.8 |
| 145 | 35.3  | 16.2  | 48.5  |

