

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
 Data File : VX035972.D
 Acq On : 02 Jun 2023 10:28
 Operator : JC/MD
 Sample : VX0602WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0602WBL01

Quant Time: Jun 02 15:50:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	209355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	337329	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	292921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127102	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	144389	42.472	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.940%
35) Dibromofluoromethane	5.391	113	104397	46.013	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.020%
50) Toluene-d8	8.647	98	393426	48.316	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.640%
62) 4-Bromofluorobenzene	11.079	95	143176	44.542	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
Data File : VX035972.D
Acq On : 02 Jun 2023 10:28
Operator : JC/MD
Sample : VX0602WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0602WBL01

Quant Time: Jun 02 15:50:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

