

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
 Data File : VX035975.D
 Acq On : 02 Jun 2023 11:43
 Operator : JC/MD
 Sample : 02958-02MEDL 10X
 Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 734MEDL

Quant Time: Jun 02 15:51:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	229929	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	371860	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	323183	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	151578	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	155758	41.717	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	83.440%		
35) Dibromofluoromethane	5.385	113	111604	44.622	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.240%		
50) Toluene-d8	8.647	98	421837	46.994	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	93.980%		
62) 4-Bromofluorobenzene	11.079	95	164220	46.344	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.680%		
Target Compounds						
16) Acetone	2.404	43	7494	4.865	ug/l #	89
18) Methyl Acetate	2.709	43	90238	15.069	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
Data File : VX035975.D
Acq On : 02 Jun 2023 11:43
Operator : JC/MD
Sample : 02958-02MEDL 10X
Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
734MEDL

Quant Time: Jun 02 15:51:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

