

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
 Data File : VX035982.D
 Acq On : 02 Jun 2023 14:31
 Operator : JC/MD
 Sample : 03022-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW11-EB-20230601

Quant Time: Jun 05 01:35:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	207383	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	332532	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	294473	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	143933	42.741	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.480%
35) Dibromofluoromethane	5.385	113	101118	45.211	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.420%
50) Toluene-d8	8.647	98	390514	48.650	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.300%
62) 4-Bromofluorobenzene	11.079	95	148650	46.912	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.820%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1356	0.976	ug/l #	83
20) Methylene Chloride	2.788	84	663	0.248	ug/l #	79
30) Chloroform	5.099	83	4674	0.984	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
Data File : VX035982.D
Acq On : 02 Jun 2023 14:31
Operator : JC/MD
Sample : 03022-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW11-EB-20230601

Quant Time: Jun 05 01:35:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

