

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
 Data File : VX035983.D
 Acq On : 02 Jun 2023 14:54
 Operator : JC/MD
 Sample : 03022-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW11-GW-518-520

Quant Time: Jun 05 01:35:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	203540	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	329105	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	286929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130916	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	142364	43.073	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.140%
35) Dibromofluoromethane	5.385	113	101303	45.765	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.540%
50) Toluene-d8	8.647	98	383350	48.255	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.500%
62) 4-Bromofluorobenzene	11.079	95	143016	45.604	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.200%
Target Compounds						
16) Acetone	2.392	43	4612	3.382	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
Data File : VX035983.D
Acq On : 02 Jun 2023 14:54
Operator : JC/MD
Sample : 03022-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW11-GW-518-520

Quant Time: Jun 05 01:35:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

