

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
 Data File : VX035984.D
 Acq On : 02 Jun 2023 15:18
 Operator : JC/MD
 Sample : 03022-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW11-GW-538-540

Quant Time: Jun 05 01:36:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	203016	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	328495	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	287105	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127865	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	143475	43.521	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.040%
35) Dibromofluoromethane	5.391	113	100921	45.677	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.360%
50) Toluene-d8	8.653	98	382042	48.180	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.360%
62) 4-Bromofluorobenzene	11.079	95	144794	46.256	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.520%
Target Compounds						
16) Acetone	2.392	43	5146	3.784	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
Data File : VX035984.D
Acq On : 02 Jun 2023 15:18
Operator : JC/MD
Sample : 03022-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW11-GW-538-540

Quant Time: Jun 05 01:36:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

