

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
 Data File : VX035986.D
 Acq On : 02 Jun 2023 16:05
 Operator : JC/MD
 Sample : 03022-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW11-DUP-20230531

Quant Time: Jun 05 01:36:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	192674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	335061	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	289609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134681	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	148495	47.462	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.920%
35) Dibromofluoromethane	5.385	113	97872	43.429	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.860%
50) Toluene-d8	8.647	98	383250	47.385	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.760%
62) 4-Bromofluorobenzene	11.079	95	148192	46.414	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.820%
Target Compounds						
16) Acetone	2.398	43	6546	5.071	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
Data File : VX035986.D
Acq On : 02 Jun 2023 16:05
Operator : JC/MD
Sample : 03022-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW11-DUP-20230531

Quant Time: Jun 05 01:36:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

