

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
 Data File : VX035987.D
 Acq On : 02 Jun 2023 16:29
 Operator : JC/MD
 Sample : 03022-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW11-GW-598-600

Quant Time: Jun 05 01:36:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	180473	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	291634	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	249877	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108938	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	128372	43.804	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	87.600%		
35) Dibromofluoromethane	5.391	113	89622	45.690	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.380%		
50) Toluene-d8	8.653	98	338678	48.109	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.220%		
62) 4-Bromofluorobenzene	11.079	95	125509	45.163	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.320%		
Target Compounds					Qvalue	
16) Acetone	2.392	43	6779	5.607	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
Data File : VX035987.D
Acq On : 02 Jun 2023 16:29
Operator : JC/MD
Sample : 03022-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW11-GW-598-600

Quant Time: Jun 05 01:36:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

