

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
 Data File : VX035988.D
 Acq On : 02 Jun 2023 16:53
 Operator : JC/MD
 Sample : 03022-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW11-GW-618-620

Quant Time: Jun 05 01:36:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	188233	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	308262	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	287862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	146621	47.968	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.940%
35) Dibromofluoromethane	5.391	113	97435	46.994	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.980%
50) Toluene-d8	8.653	98	384541	51.678	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.360%
62) 4-Bromofluorobenzene	11.079	95	146025	49.712	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.420%
Target Compounds						
16) Acetone	2.398	43	8263	6.553	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
Data File : VX035988.D
Acq On : 02 Jun 2023 16:53
Operator : JC/MD
Sample : 03022-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW11-GW-618-620

Quant Time: Jun 05 01:36:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

