

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
 Data File : VX035990.D
 Acq On : 02 Jun 2023 17:40
 Operator : JC/MD
 Sample : 03022-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW11-GW-658-660

Quant Time: Jun 05 09:32:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	181989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	315069	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	278875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127994	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	143541	48.572	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.140%
35) Dibromofluoromethane	5.391	113	93586	44.162	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.320%
50) Toluene-d8	8.653	98	372283	48.950	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.900%
62) 4-Bromofluorobenzene	11.079	95	140846	46.913	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.820%
Target Compounds						
16) Acetone	2.385	43	2684	2.201	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060223\
Data File : VX035990.D
Acq On : 02 Jun 2023 17:40
Operator : JC/MD
Sample : 03022-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW11-GW-658-660

Quant Time: Jun 05 09:32:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

