

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060225\
 Data File : VX046450.D
 Acq On : 02 Jun 2025 17:32
 Operator : JC/MD
 Sample : Q2161-01
 Misc : 5.16g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B27-SOIL-SAMPLE

Quant Time: Jun 03 01:34:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	63815	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124373	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	117389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54091	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59721	50.198	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.400%
35) Dibromofluoromethane	5.379	113	42427	47.372	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.740%
50) Toluene-d8	8.647	98	156991	50.645	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.280%
62) 4-Bromofluorobenzene	11.079	95	61868	52.031	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060225\
Data File : VX046450.D
Acq On : 02 Jun 2025 17:32
Operator : JC/MD
Sample : Q2161-01
Misc : 5.16g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B27-SOIL-SAMPLE

Quant Time: Jun 03 01:34:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

